

## Quantification of Human Cytomegalovirus DNA by Real-Time PCR

ELYANNE GAULT,<sup>1,2†</sup> YANNE MICHEL,<sup>1</sup> AXELLE DEHÉE,<sup>1,2</sup> CHAHRAZED BELABANI,<sup>1</sup>  
JEAN-CLAUDE NICOLAS,<sup>2,3</sup> AND ANTOINE GARBARG-CHENON<sup>1,2\*</sup>

*Laboratoire de Virologie, Hôpital Armand Trousseau,<sup>1</sup> Laboratoire de Virologie, Hôpital Rothschild,<sup>3</sup>  
and EA 2391, UFR Saint-Antoine,<sup>2</sup> Paris, France*

Received 28 September 2000/Returned for modification 23 October 2000/Accepted 27 November 2000

**A quantitative real-time PCR assay was developed to measure human cytomegalovirus (HCMV) DNA load in peripheral blood leukocytes (PBLs). The HCMV DNA load in PBLs was normalized by means of the quantification of a cellular gene (albumin). The results of the real-time PCR assay correlated with those of the HCMV pp65-antigenemia assay ( $P < 0.0001$ ).**

Human cytomegalovirus (HCMV) infection is characterized by a primary infection leading to a lifelong persistence of the viral genome. Periodically, the virus reactivates from latency and recovers its ability to multiply. HCMV is a major cause of morbidity and mortality in bone marrow or solid-organ transplant recipients and in AIDS patients. Early diagnosis of HCMV infection in high-risk patients is essential in order to start preemptive treatments (3, 13, 15). The detection of the pp65 antigen in leukocytes is a sensitive method widely used for the early diagnosis of HCMV infection, but it is labor-intensive, requires immediate processing, and relies on a subjective interpretation of the slides (1, 7). Qualitative PCR detection of HCMV DNA in leukocytes or plasma is considered the most sensitive method, but it lacks specificity for the diagnosis of HCMV disease (2, 6, 16). Quantification of HCMV DNA has been proposed to be more specifically associated with the disease (1, 18). Real-time PCR based on the TaqMan technology (4, 5) provides an accurate means to quantify viral DNA with the major advantage of avoiding post-PCR handling that can be the source of DNA carryover. Recent studies report the utility of real-time PCR for the quantification of HCMV DNA (9, 11, 12, 14, 17). In these studies, the primers used for PCR were located in the major immediate-early gene (12, 14, 17), the US17 gene (9), or the glycoprotein B gene (17). As suggested by Yun et al., the sensitivity of quantitative PCR may be dependent on the viral target gene (17); however, the most appropriate region for amplification has not been established, as the sequence diversity of clinical isolates remains to be characterized. In the present study, a real-time PCR assay was developed to quantify HCMV DNA in peripheral blood leukocytes (PBLs) using a target sequence located in the UL83 gene which codes for the lower matrix protein detected in the pp65 antigenemia test. The HCMV DNA load in PBLs was normalized by means of the quantification of a

cellular gene (albumin) and the results were compared to those of the pp65 antigenemia assay.

DNA extractions were performed in all experiments using the QIAamp blood kit (QIAGEN S. A., Courtabœuf, France) according to the manufacturer's recommendations, except that DNA was eluted in 200  $\mu$ l of distilled water. To amplify HCMV DNA, primers and probe were defined in the UL83 region as follows: pp549s (direct primer), 5'-GTCAGCGTTC GTGTTTCCCA-3'; pp812as (reverse primer), 5'-GGGACAC AACACCGTAAAGC-3'; and pp770s (fluorogenic probe), 5'-FAM-CCC GCAACCCGCAACCCCTTCATG-3'-TAMRA. No cross-reactivity was observed when the specificity of the primers and probe was tested for other human herpes viruses (herpes simplex virus types 1 and 2, varicella-zoster virus, human herpes virus 6, and Epstein-Barr virus) and human fibroblast DNA (data not shown). The real-time PCRs were carried out using the TaqMan PCR core reagent kit (PE Applied Biosystems, Courtabœuf, France). Ten microliters of DNA was added to a PCR mixture containing 1 $\times$  PCR buffer (10 mM Tris-HCl [pH 8.3], 50 mM KCl, 6-carboxy-x-rhodamine), 5 mM MgCl<sub>2</sub>, 0.2 mM dATP, dCTP, and dGTP, 0.4 mM dUTP, 0.2 mM concentrations of each primer, 0.1 mM fluorogenic probe, 1.25 U of AmpliTaq Gold, and 0.5 U of AmpErase. The PCR conditions consisted of 1 cycle of 2 min at 50°C and 1 cycle of 10 min at 95°C followed by 45 cycles of 15 s at 95°C and 1 min at 65°C. The reaction, data acquisition, and analysis were performed using the ABI PRISM 7700 sequence detection system (PE Applied Biosystems). The number of target copies in the reaction was deduced from the threshold cycle ( $C_T$ ) values corresponding to the fractional cycle number at which the released fluorescence exceeded 20 times the standard deviation of the mean baseline emission. A plasmid containing one copy of the UL83 target sequence (pKS-pp65K7) was used as a standard for HCMV DNA quantification. To construct the pKS-pp65K7 standard, DNA from the HCMV AD169 strain was amplified with primers pp549s and pp812as and cloned into the pBluescript II KS vector (Stratagene, Amsterdam, The Netherlands). The concentration of purified pKS-pp65K7 plasmid DNA was determined with a spectrophotometer at 260 nm and the corresponding copy number was calculated.

To evaluate HCMV DNA quantification on clinical samples, 46 PBL samples that were positive in the pp65 antigenemia

\* Corresponding author. Mailing address: Laboratoire de Virologie, Hôpital Armand Trousseau, 26 Avenue du Dr. Arnold Netter 75571, Paris Cedex 12, France. Phone: 33 1 44 73 62 81. Fax: 33 1 44 73 62 88. E-mail: a.chenon@trs.ap-hop-paris.fr.

† Present address: Laboratoire de Virologie, Hôpital Avicenne, 93009 Bobigny cedex.

assay were collected from solid-organ transplant recipients or AIDS patients. The pp65 antigenemia assay was carried out on fresh PBL samples using the CINAKit (Argene Biosoft, Varilhes, France) following the manufacturer's instructions. DNA was extracted from  $2 \times 10^6$  PBLs isolated from EDTA-treated blood samples. For each sample, 2 aliquots of DNA, corresponding to  $10^5$  cells, were used to quantify HCMV DNA using the standard curve constructed with pKS-pp65K7. Two other aliquots were used to quantify the human albumin gene using primers and probe described by Laurendeau et al. (8) with some modifications: the forward primer was 5'-GCTGTCATCTCTTG TGGGCTGT-3', the reverse primer was 5'-AAACTCATGGG AGCTGCTGGTT-3', and the probe was 5'FAM-CCTGTCAT GCCCACAAATCTCTCC-3' TAMRA. The conditions used to amplify the albumin gene were identical to those used for HCMV DNA, allowing both reactions to be carried out simultaneously in the same run. The standard curve for quantification of the albumin gene was constructed with 10-fold serial dilutions ranging from  $2 \times 10^5$  to 20 copies of human genomic DNA (Roche Molecular Biochemicals, Meylan, France) tested in triplicate. The normalized value of the HCMV DNA load was expressed as the number of HCMV copies per  $2 \times 10^5$  PBLs calculated as the ratio (HCMV DNA mean copy number/[albumin DNA mean copy number/2])  $\times 2 \times 10^5$ . This allowed sample-to-sample variations, which can lead to a misinterpretation of the HCMV DNA copy number, to be normalized in terms of cell count, yield of DNA extraction, and PCR efficiency.

The sensitivity and the intra-assay reproducibility of the real-time PCR assay were evaluated by using plasmid pKS-pp65K7 (containing one copy of the UL83 target sequence) as a standard. One copy of the plasmid could be detected in one in five assays, while 10 plasmid DNA copies were detected with 100% sensitivity. Thus, the lowest limit for quantification was considered to be 10 copies of the target sequence. The intra-assay variation of the  $C_T$  values was evaluated using five replicates of plasmid dilutions containing  $10^6$ ,  $10^5$ ,  $10^4$ ,  $10^3$ ,  $10^2$ , and 10 copies. The coefficients of variation (CV) were 1.4, 1.9, 0.6, 1.3, 1.1, and 2.5%, respectively. Plasmid pKS-pp65K7 was used to construct the standard curve for HCMV DNA quantification. Tenfold serial dilutions ranging from  $10^6$  to 10 copies of plasmid were tested in triplicate and the  $C_T$  values were plotted against the copy number. The linear correlation between the  $C_T$  and the logarithm of the DNA copy number was repeatedly greater than 0.995 (Fig. 1A).

To determine whether the results of HCMV DNA quantification by real-time PCR were consistent with those obtained by virus titration in a shell vial assay, serial dilutions of a viral stock of AD169 with a titer of  $7 \times 10^6$  infectious foci (FFU)/ml were quantified with the TaqMan assay (Fig. 1B). One FFU corresponded to about 20 HCMV DNA copies. Moreover, as shown in Fig. 1, the DNA amplification efficiencies obtained with AD169 and pKS-pp65K7 were similar (95 and 98%, respectively). This result gave support to the use of plasmid pKS-pp65K7 as a relevant standard for HCMV DNA quantification. The interassay reproducibility of the TaqMan assay was evaluated using 3 serial dilutions of the AD169 stock ( $10^3$ ,  $10^2$ , and 10 FFU). For each dilution, four independent DNA extractions were performed and HCMV DNA was quantified in four independent TaqMan runs. The CV values of the




FIG. 1. Standard curves for HCMV DNA quantification. (A) Plasmid pKS-pp65K7 containing one copy of the UL83 target sequence was used to construct the standard curve for HCMV DNA quantification. Tenfold serial dilutions ranging from  $10^6$  to 10 copies of plasmid were tested in triplicate and the mean  $C_T$  values were plotted against the copy number. The correlation coefficient was 0.997 and the slope was  $-3.36$ . The amplification efficiency, calculated as  $[10^{(-1/\text{slope})} - 1] \times 100$ , was 98%. (B) Serial dilutions of HCMV strain AD169 (obtained from a cell culture supernatant with a titer of  $7 \times 10^6$  FFU/ml in a shell vial assay) were tested by real-time PCR and the mean  $C_T$  values were plotted against the number of FFU input in the reaction mixture. The correlation coefficient was 0.999. The slope was  $-3.46$ , corresponding to 95% amplification efficiency.

HCMV DNA copy numbers were 17, 12, and 21%, respectively, for the 3 dilutions tested.

To evaluate the HCMV DNA TaqMan assay on clinical samples, we tested 46 PBL samples that were positive in the pp65 antigenemia assay. HCMV DNA quantification in PBL samples was performed in parallel with the quantification of the albumin gene in order to determine the amount of cellular DNA input in each reaction. The mean value of the albumin gene copy number for the 46 PBL samples tested was  $6.2 \times 10^4$  copies (ranging from  $7.2 \times 10^3$  to  $2 \times 10^5$  copies). This corresponded to a mean DNA extraction yield of 31%, compared to the theoretical value of  $2 \times 10^5$  albumin copies expected for the amount of cellular DNA input in the reaction mixture (equivalent to  $10^5$  cells). Forty-four out of 46 PBL samples were positive for HCMV DNA with the TaqMan assay. For the 44 HCMV DNA-positive samples, the normalized values of the HCMV DNA load ranged from 1 to  $4.26 \times 10^5$  copies per  $2 \times 10^5$  cells. As shown in Fig. 2, a statistically significant correlation was observed between the HCMV DNA copy number and the pp65-positive cell number (Spearman rank test;  $r = 0.657$ ,  $P < 0.0001$ ). As the level of HCMV antigenemia is usually considered to correlate with HCMV disease (1), samples from patients were classified into three groups according to the




FIG. 2. Correlation between the HCMV DNA copy number and the pp65-positive cell number in PBL. HCMV DNA load was quantified by real-time PCR in 46 PBL samples that were positive in the pp65 antigenemia assay. The copy numbers of HCMV DNA were normalized by means of the quantification of the albumin gene and plotted against the pp65-positive cell numbers obtained by the antigenemia assay (pp65 values greater than 200 positive cells per  $2 \times 10^5$  PBLs were set to 200). The two samples negative for HCMV DNA were not figured on this logarithmic graph. The correlation between the HCMV DNA copy number and the pp65-positive cell number was examined by the nonparametric Spearman rank test and was found to be significant, with a correlation coefficient of  $r = 0.657$  ( $P < 0.0001$ ).

results of the pp65 antigenemia assay. Group 1 ( $n = 26$ ) corresponded to samples with low HCMV antigenemia values ( $<10$  positive cells/ $2 \times 10^5$  PBLs), group 2 ( $n = 12$ ) corresponded to moderate values (10 to 50 positive cells/ $2 \times 10^5$  PBLs), and group 3 ( $n = 8$ ) corresponded to high values ( $\geq 100$  positive cells/ $2 \times 10^5$  PBLs). As shown in Fig. 3, the high pp65-positive samples from group 3 had significantly higher HCMV DNA copy numbers than the moderate pp65-positive samples from group 2 (Mann-Whitney U test;  $P = 0.0002$ ) and the low pp65-positive samples from group 1 ( $P < 0.0001$ ). The HCMV DNA load values were also significantly different between groups 1 and 2 ( $P = 0.0009$ ). The mean HCMV DNA copy number per  $2 \times 10^5$  PBLs was  $1.8 \times 10^2$  (range, 0 to  $3.5 \times 10^3$ ) for group 1 samples,  $5.8 \times 10^2$  (range, 20 to  $2 \times 10^3$ ) for group 2, and  $1.3 \times 10^5$  (range,  $8.7 \times 10^3$  to  $4.6 \times 10^6$ ) for group 3.

The development of real-time PCR technology is a promising improvement for the quantification of HCMV DNA in clinical samples and will be useful for the follow-up of patients with a high risk of developing HCMV disease. The real-time PCR assay developed in this study allowed the quantification of HCMV DNA over a large dynamic range (10 to  $10^6$  copies), as previously reported for other real-time PCR assays (9, 11, 12, 14, 17). The assay was reproducible as indicated by the intra-assay CV values obtained with the plasmid standard (0.6 to 2.5%) and by the interassay CV values obtained with the HCMV AD169 strain (12 to 21%). The fact that the evaluation of interassay reproducibility took into account the DNA extraction step could explain why the CV values were higher than those ( $<10\%$ ) reported by Nitsche et al. with a quantitative HCMV DNA TaqMan-based assay (12). Because only one PBL sample was available for each patient, replicates of DNA extraction and quantification could not be performed in order




FIG. 3. Comparison of HCMV DNA load with pp65 antigenemia levels in PBL samples. The HCMV DNA load was quantified by real-time PCR in 46 PBL samples and compared to the pp65 antigenemia level. High pp65-positive samples ( $\geq 100$  positive cells/ $2 \times 10^5$  PBLs;  $n = 8$ ) had significantly higher HCMV DNA copy numbers than the moderate pp65-positive samples (10 to 50 positive cells;  $n = 12$ ) and the low pp65-positive samples ( $<10$  positive cells;  $n = 26$ ).  $p$  indicates the significance of the nonparametric Mann-Whitney U test. Bars show the mean HCMV DNA copy number and standard error of the mean for each group.

to determine the CV values for clinical specimens. For such specimens the DNA extraction step could have a major effect on the assay's reproducibility. Indeed, the quantification of the albumin gene in PBL samples showed a wide range of sample-to-sample variation in the DNA extraction yield, emphasizing the need to normalize the HCMV load values for cell samples as proposed by others (10). This was achieved by referring the HCMV copy number not to the initial cell count but to the actual amount of cellular DNA estimated by quantification of the albumin gene.

The HCMV DNA copy number in PBL samples determined with real-time PCR correlated to the pp65-positive cell number, in agreement with the results of other TaqMan-based assays (9, 14). However, some discrepancies were observed. First, two samples with low antigenemia pp65 values (2 and 4 positive cells/ $2 \times 10^5$  PBLs) were negative for HCMV DNA. For both samples, the albumin gene quantification indicated a poor DNA extraction yield ( $\leq 10\%$ ). Such cases of low DNA yield should lead to cautious interpretation of negative HCMV DNA results. It was not possible to retest these two samples due to insufficient material. Secondly, although 21 out of 26 samples from group 1 (pp65 antigenemia was  $<10$  positive cells/ $2 \times 10^5$  PBLs) had an HCMV copy number per  $2 \times 10^5$  PBLs of less than 100, 4 samples from this group ranged from 101 to 372 copies, and 1 sample had  $3.5 \times 10^3$  copies. Others have reported a similar difference between antigenemia and quantitative PCR results (10, 16), but the clinical significance of a high HCMV DNA load associated with a low antigenemia remains to be established.

As for other quantitative PCR techniques, real-time PCR

applied to the quantification of HCMV DNA needs to be standardized in order to ameliorate the reproducibility and to ensure the most accurate follow-up of the patients. The region of the viral genome which is the most appropriate target for HCMV quantitative PCR remains to be determined, and a comparative evaluation of TaqMan assays based on different HCMV target genes will be useful for this purpose. Moreover, the standardization of the technique requires the amplification of a control (such as a cellular gene) to monitor the efficiency of the reaction. As the TaqMan technology allows multiplex PCR to be performed by using two fluorogenic probes labeled with different dyes, the dual quantification of HCMV DNA and the albumin gene in a single-tube format will be the next step in the development of our assay.

This work was supported in part by an ANRS (AC 11, groupe CMV) grant and by the MESRT grant Programme de Recherches Fondamentales en Microbiologie, Maladies infectieuses et Parasitologie.

We thank Ann Beaumont for careful reading and editing of the manuscript.

#### REFERENCES

1. Boeckh, M., and G. Boivin. 1998. Quantitation of cytomegalovirus: methodologic aspects and clinical applications. *Clin. Microbiol. Rev.* **11**:533–554.
2. Drouet, E., R. Colimon, S. Michelson, N. Fourcade, A. Niveleau, C. Ducerf, A. Boibieux, M. Chevallier, and G. Denoyel. 1995. Monitoring levels of human cytomegalovirus DNA in blood after liver transplantation. *J. Clin. Microbiol.* **33**:389–394.
3. Einsele, H., G. Ehninger, H. Hebart, K. M. Wittkowski, U. Schuler, G. Jahn, P. Mackes, M. Herter, T. Klingebiel, J. Löffler, S. Wagner, and C. A. Müller. 1995. Polymerase chain reaction monitoring reduces the incidence of cytomegalovirus disease and the duration and side effects of antiviral therapy after bone marrow transplantation. *Blood* **86**:2815–2820.
4. Heid, C. A., J. Stevens, K. J. Livak, and P. M. Williams. 1996. Real time quantitative PCR. *Genome Res.* **6**:986–994.
5. Holland, P. M., R. D. Abramson, R. Watson, and D. H. Gelfand. 1991. Detection of specific polymerase chain reaction product by utilizing the 5'-3' exonuclease activity of *Thermus aquaticus* DNA polymerase. *Proc. Natl. Acad. Sci. USA* **88**:7276–7280.
6. Kanj, S. S., A. I. Sharara, P. A. Clavien, and J. D. Hamilton. 1996. Cytomegalovirus infection following liver transplantation: review of the literature. *Clin. Infect. Dis.* **22**:537–549.
7. Landry, M. L., D. Ferguson, S. Cohen, K. Huber, and P. Wetherill. 1995. Effect of delayed specimen processing on cytomegalovirus antigenemia test results. *J. Clin. Microbiol.* **33**:257–259.
8. Laurendeau, I., M. Bahuaud, N. Vodovar, C. Larramendy, M. Olivi, I. Bieche, M. Vidaud, and D. Vidaud. 1999. TaqMan PCR-based gene dosage assay for predictive testing in individuals from a cancer family with INK4 locus haploinsufficiency. *Clin. Chem.* **45**:982–986.
9. Machida, U., M. Kami, T. Fukui, Y. Kazuyama, M. Kinoshita, Y. Tanaka, Y. Kanda, S. Ogawa, H. Honda, S. Chiba, K. Mitani, Y. Muto, K. Osumi, S. Kimura, and H. Hirai. 2000. Real-time automated PCR for early diagnosis and monitoring of cytomegalovirus infection after bone marrow transplantation. *J. Clin. Microbiol.* **38**:2536–2542.
10. Nazzari, C., A. Gaeta, M. Lazzarini, T. D. Castelli, and C. Mancini. 2000. Multiplex polymerase chain reaction for the evaluation of cytomegalovirus DNA load in organ transplant recipients. *J. Med. Virol.* **61**:251–258.
11. Nitsche, A., N. Steuer, C. A. Schmidt, O. Landt, H. Ellerbrok, G. Pauli, and W. Siebert. 2000. Detection of human cytomegalovirus DNA by real-time quantitative PCR. *J. Clin. Microbiol.* **38**:2734–2737.
12. Nitsche, A., N. Steuer, C. A. Schmidt, O. Landt, and W. Siebert. 1999. Different real-time PCR formats compared for the quantitative detection of human cytomegalovirus DNA. *Clin. Chem.* **45**:1932–1937.
13. Roberts, T. C., D. C. Brennan, R. S. Buller, M. Gaudreault-Keener, M. A. Schnitzler, K. E. Sternhell, K. A. Garlock, G. G. Singer, and G. A. Storch. 1998. Quantitative polymerase chain reaction to predict occurrence of symptomatic cytomegalovirus infection and assess response to ganciclovir therapy in renal transplant recipients. *J. Infect. Dis.* **178**:626–635.
14. Tanaka, N., H. Kimura, K. Iida, Y. Saito, I. Tsuge, A. Yoshimi, T. Matsuyama, and T. Morishima. 2000. Quantitative analysis of cytomegalovirus load using a real-time PCR assay. *J. Med. Virol.* **60**:455–462.
15. van der Meer, J. T., W. L. Drew, R. A. Bowden, G. J. Galasso, P. D. Griffiths, D. A. Jabs, C. Katlama, S. A. Spector, and R. J. Whitley. 1996. Summary of the international consensus symposium on advances in the diagnosis, treatment and prophylaxis and cytomegalovirus infection. *Antiviral Res.* **32**:119–140.
16. Weber, B., U. Nestler, W. Ernst, H. Rabenau, J. Braner, A. Birkenbach, E. H. Scheuermann, W. Schoeppe, and H. W. Doerr. 1994. Low correlation of human cytomegalovirus DNA amplification by polymerase chain reaction with cytomegalovirus disease in organ transplant recipients. *J. Med. Virol.* **43**:187–193.
17. Yun, Z., I. Lewensohn-Fuchs, P. Ljungman, and A. Vahlne. 2000. Real-time monitoring of cytomegalovirus infections after stem cell transplantation using the TaqMan polymerase chain reaction assays. *Transplantation* **69**:1733–1736.
18. Zipeto, D., F. Baldanti, D. Zella, M. Furione, A. Cavicchini, G. Milanese, and G. Gerna. 1993. Quantification of human cytomegalovirus DNA in peripheral blood polymorphonuclear leukocytes of immunocompromised patients by the polymerase chain reaction. *J. Virol. Methods* **44**:45–55.