

False-Positive Results for Rapid Diagnostic Tests for Malaria in Patients with Rheumatoid Factor

Jong-Han Lee,^a Jin Woo Jang,^a Chi Hun Cho,^a Ju Yeon Kim,^a Eun Taek Han,^b Seung Gyu Yun,^c Chae Seung Lim^a

Department of Laboratory Medicine, College of Medicine, Korea University, Seoul, South Korea^a; Department of Parasitology, College of Medicine, Kangwon National University, Chuncheon, South Korea^b; The Armed Forces Medical School, Daejeon, South Korea^c

Four different rapid diagnostic tests (RDTs) for malaria were evaluated by testing 82 healthy control patients, 89 *Plasmodium vivax*-infected patients, and 92 rheumatoid factor (RF)-positive nonmalaria patients. The false-positive rate ranged from 2.2% to 13% in RF-positive patients. High RF levels are associated with malaria RDT false positivity.

Malaria remains a major global health problem in tropical and subtropical countries, with high morbidity and mortality and extensive economic loss (1). Malaria rapid diagnostic tests (RDTs) are becoming the clinical diagnostic method of choice due to their quick results and ease of use, even by inexperienced personnel (2). However, false-positive results may be observed in patients with rheumatoid factor (RF), hepatitis C, toxoplasmosis, human African trypanosomiasis, dengue, leishmaniasis, Chagas disease, and schistosomiasis (2). Iqbal et al. (3) reported that 33 of the 35 false-positive specimens were negative when the RF was absorbed in the immunochromatographic test (ICT). The goal of this study was to use four different malaria RDTs to explore the relationship between false-positive malaria RDT results and RF.

Between April 2010 and August 2013, a total of 263 whole-blood samples with EDTA were collected from South Korean patients at the Korea University Guro Hospital, Republic of Korea. Of these 263 samples, 89 were infected with malaria, as confirmed by Giemsa-stained microscopic examination, 92 did not have malaria but did have RF, and 82 had neither malaria nor RF. Both microscopy and PCR were used to rule out malaria. Each patient provided informed consent under the protocol for human use, which was approved by the Human Use Ethical Committee, Korea University Guro Hospital.

Thick and thin blood films were prepared when blood was drawn in accordance with standard procedures. These films were stained with Giemsa and examined by trained microscopists who did not have prior knowledge of the patients' clinical history. *Plasmodium* species and the parasite density were determined. The circumsporozoite protein (CSP) gene of *Plasmodium vivax* was amplified by PCR using previously established methods (4). Four commercial malaria RDT kits were selected based on the multiple target antigens (Ags) detected, the BinaxNOW malaria kit (Binax Inc., Scarborough, ME, USA), the OptiMAL-IT malaria kit (Bio-Rad, Marnes la Coquette, France), the SD Bioline malaria Ag Pf/Pan rapid test (Standard Diagnostics, Inc., Yongin, South Korea), and the Humasis malaria P.f/Pan antigen test (Humasis, Anyang, South Korea). BinaxNOW detects both histidine-rich protein 2 (HRP-2), which is specific to *Plasmodium falciparum*, and aldolase, which is a pan-malarial enzyme found in the five human pathogenic *Plasmodium* species (5). OptiMAL-IT differentiates *P. falciparum*-specific lactate dehydrogenase (pLDH) and pan-*Plasmodium* lactate dehydrogenase (pLDH) by immunological detection (6). The SD Bioline and Humasis tests target HRP-2 for *P. falciparum* and pLDH for other human malaria species (7, 8). All

tests were performed according to the manufacturers' instructions. The RDT results were interpreted by multiple technicians.

RF quantitation was performed using a Seiken RF latex(X1) reagent (Denka Seiken, Tokyo, Japan) in a TBA-200FR NEO automatic analyzer (Toshiba Medical Systems Corporation, Tochigi-ken, Japan). We tested the specimens for antinuclear antibody (ANA) using the Kallestad HEp-2 cell line substrate (Bio-Rad Laboratories, Hercules, CA, USA), according to the manufacturer's instructions, in order to discriminate the ANA effect on false-positive malaria RDTs. We used two enzyme-linked immunosorbent assays (ELISAs) that target different malaria antigens for comparison of malaria RDTs, the SD malaria antigen Pf ELISA (Standard Diagnostics, Inc.) and the Genedia malaria antigen ELISA (Green Cross Co., Seoul, South Korea). The SD malaria antigen Pf ELISA and Genedia malaria antigen ELISA were used to detect HRP-2 and pLDH, respectively.

No malaria RDT showed false-positive results in *P. falciparum*-specific bands from the 89 patients infected with *P. vivax* and the 82 healthy controls. Of the 92 RF-positive specimens, there were 15 false positives (16.3%; 95% confidence interval, 0.1013 to 0.2517), and of the 368 (92 × 4) malaria RDT results, there were 26 false positives (7.1%; 95% confidence interval, 0.0487 to 0.1016) (Table 1). BinaxNOW had the highest false-positive rate by specimen (13%), with a rate of 9.8% for the HRP-2 and 5.4% for the aldolase bands. The SD Bioline test had the lowest false-positive rate by specimen (2.2%), with a rate of 1.1% for the HRP-2 and 1.1% for the pLDH bands. The Humasis test and the OptiMAL-IT test had a 6.5% false-positive rate by specimen. The mean RF levels were 3.2 ± 2.8 IU/ml (range, 1.0 to 14.8 IU/ml) in the healthy control patients (n = 82), 6.4 ± 5.2 IU/ml (range, 1.0 to 24.8 IU/ml) in the *P. vivax*-positive patients (n = 89), and 270.8 ± 299.2 IU/ml (range, 16.2 to 1,452.1 IU/ml) in the RF-positive patients (n = 92). The mean levels of RF were lowest

Received 27 June 2014 Returned for modification 30 June 2014

Accepted 18 July 2014

Published ahead of print 23 July 2014

Editor: M. J. Loeffelholz

Address correspondence to Chae Seung Lim, malarim@korea.ac.kr.

J.-H.L. and J.W.J. contributed equally to this article.

Copyright © 2014, American Society for Microbiology. All Rights Reserved.

doi:10.1128/JCM.01797-14

TABLE 1 Characteristics associated with 15 false-positive results produced by malaria RDT kits and malaria ELISAs

No. of RDTs used	Sample no.	Gender ^a	Age (yr)	IgM RFP ^b (IU/ml)	Test results ^c							Absorbance value of SD Bioline malaria antigen Pf detection ^d (cutoff, 0.1575)	Absorbance value of Genedia malaria antigen ELISA for PLDH detection ^d (cutoff, 0.12)	ANA ^e Diagnosis ^f		
					BinaxNOW malaria HRP-2	Aldolase	OptiMAL-IT PLDH	PLDH	PLDH	HRP-2	LDH				HRP-2	LDH
1 (RF: 348 ± 277.7 IU/ml)	1	F	36	113.9	+	(weak)	-	-	-	-	-	-	-	-	-	Rheumatoid arthritis
	2	F	73	117.4	-	-	-	-	-	+	(weak)	-	-	-	-	HTN DM
	3	M	54	156.1	+	(weak)	-	-	-	-	-	-	-	-	-	Rheumatoid arthritis
	4	F	20	208.9	+	(weak)	-	-	-	-	-	-	-	-	-	Arthritis
	5	M	56	338.8	+	-	-	-	-	-	-	-	-	-	-	Asthma
	6	M	44	410.1	+	(weak)	-	-	-	-	-	-	-	-	-	Health check-up
	7	F	18	500.3	-	-	+	(weak)	+	(weak)	-	-	-	-	-	Rheumatoid arthritis
	8	M	62	938.5	-	-	-	-	-	-	-	+	(weak)	+	(weak)	Rheumatoid arthritis
2 (RF: 713.4 ± 608.8 IU/ml)	9	M	45	341.8	+	(weak)	-	-	-	-	-	+	(weak)	-	-	Health check-up
	10	F	71	382.5	+	(weak)	-	-	-	-	-	+	(weak)	-	-	Bursitis
	11	F	43	1,416.0	+	(weak)	+	(weak)	+	(weak)	-	-	-	-	-	HBV
3 (RF: 1147.5 ± 292.0 IU/ml)	12	F	33	817.6	+	(weak)	+	+	+	+	+	+	(weak)	-	-	HBV
	13	F	53	995.9	-	-	+	(weak)	+	(weak)	-	-	-	-	-	Rheumatoid arthritis
	14	M	47	1,324.2	-	-	+	(weak)	+	(weak)	-	-	-	-	-	Rheumatoid arthritis
	15	M	55	1,452.1	-	-	+	(weak)	+	(weak)	-	-	-	-	-	Arthrosis

^a F, female; M, male.
^b RF, rheumatoid factor.
^c HRP-2, histidine-rich protein 2; pLDH, *P. falciparum* lactate dehydrogenase; pLDH, pan-*Plasmodium* lactate dehydrogenase; +, positive; -, negative.
^d Enzyme-linked immunosorbent assay (ELISA)-positive results greater than the cutoff value are noted in bold type.
^e RF levels in the antinuclear antibody (ANA)-positive group were not significantly different from those of the ANA-negative group (Mann-Whitney U test, $P = 0.346$).
^f HTN, hypertension; DM, diabetes mellitus; HBV, hepatitis B virus infection.




FIG 1 Rheumatoid factor (RF) levels according to number of reactive malaria rapid diagnostic tests (RDTs). The control group shows RF levels for negative RDT results. The upper and lower boxes show interquartile RF values, the inner lines indicate median RF values, and the whiskers illustrate the standard deviation around the mean.

in the samples with a single positive malaria RDT result (348 ± 277.7 IU/ml) and highest in the cases with three positive RDT results ($1,147.5 \pm 292.0$ IU/ml) (Fig. 1). The HRP-2-based SD ELISA had a false-positive rate of 67.4% (62/92) in RF-positive specimens, with a mean RF level of 332.1 ± 308.8 IU/ml (range, 101.7 to 1,452.1 IU/ml), while the pLDH-based Genedia ELISA had a false-positive rate of 33.7%, with a mean RF level of 288.3 ± 179.1 IU/ml (range, 101.7 to 938.5 IU/ml). Of the 15 cases of false positives in the RF-positive group, the antinuclear antibody was present in 9 of the serum samples (60%).

Many types of malaria RDTs are available, and the World Health Organization (WHO) and Foundation for Innovative New Diagnostics (FIND) have carried out detailed performance assessments of these tests (9–12). The exact mechanism behind the reaction of RF with malaria RDTs has not been fully elucidated. However, one possible explanation for the false positives observed in malaria RDTs is that there is a reaction between RF and specific antibodies on the malaria RDT strips (13–15). RF is an autoantibody directed against antigenic determinants on the fragment crystallizable (Fc) region of immunoglobulin G (IgG) molecules; thus, RFs can bind to the trapping antibody (colloidal gold-labeled antibody). This antibody complex is easily detected by capturing the antibody on the strip, resulting in a false positive (14). This study systematically investigated the pattern of false positives in malaria RDTs. The composition of the target antigen varies between malaria RDTs. HRP-2-based RDTs are more sensitive than aldolase- or pLDH-based RDTs for *P. falciparum*, whereas pLDH-based RDTs are more sensitive than aldolase-based RDTs for *P. vivax* (2, 10). Nevertheless, specificity is also important for selecting malaria RDTs, and our study suggests new points for consideration for malaria RDTs to reduce false-positive results. In addition, the prevalence of rheumatoid arthritis (RA) is about 1% in Caucasians but ranges from 0.1% in rural Africans to 5% in indigenous peoples of North America (16, 17). This difference could be related to the structural characteristics of RF. Soltys et al. (18) proposed that some RFs may not bind to certain types of IgG or that not all IgG can be bound by RFs. The highest false-positive rate was for the anti-HRP-2 band, which is specific to *P. falciparum*.

This finding was also observed in malaria ELISA results, with a false-positive rate of 67.4% and 33.7% for HRP-2 and pLDH-based malaria Ag ELISA, respectively. Suspected malaria cases that are HRP-2 positive without microscopy should undergo confirmation tests, including malaria RDTs that target other malaria antigens such as pLDH and aldolase. Compared to the pLDH or aldolase band, the HRP-2 band has a higher rate of false positives at lower levels of RF. Malaria RDT false positives tend to increase with increased RF levels. These findings were identified in all four types of malaria RDTs. Of the 62 false-positive HRP-2 ELISA results, 31 were positive for antinuclear antibody (ANA), and 32 were negative for ANA. Of the 31 false-positive pLDH ELISA results, 12 were ANA positive, and 19 were ANA negative. RF levels in the ANA-positive group were not significantly different from those in the ANA-negative group. There was no clear false-positive pattern according to ANA presence or absence. High RF levels may be able to produce false-positive malaria RDT results regardless of ANA presence. High RF levels tend to be associated with false positives in malaria RDTs. We should carefully interpret malaria RDTs and consider clinical situations and laboratory, and especially microscopic confirmation tests in cases of suspected false-positive results for malaria RDT.

ACKNOWLEDGMENTS

This research was supported by National Research Foundation of Korea grant 2012R1A1A2006650, funded by the South Korean government.

We have no conflicts of interest to declare.

REFERENCES

- World Health Organization. 2008. Global malaria control and elimination: report of a technical review. World Health Organization, Geneva, Switzerland.
- Maltha J, Gillet P, Jacobs J. 2013. Malaria rapid diagnostic tests in endemic settings. *Clin. Microbiol. Infect.* 19:399–407. <http://dx.doi.org/10.1111/1469-0691.12151>.
- Iqbal J, Khalid N, Hira PR. 2003. Performance of rapid malaria Pf antigen test for the diagnosis of malaria and false-reactivity with autoantibodies. *Adv. Exp. Med. Biol.* 531:135–148. http://dx.doi.org/10.1007/978-1-4615-0059-9_10.
- Oh JS, Kim JS, Lee CH, Nam DH, Kim SH, Park DW, Lee CK, Lim CS, Park GH. 2008. Evaluation of a malaria antibody enzyme immunoassay for use in blood screening. *Mem. Inst. Oswaldo Cruz* 103:75–78. <http://dx.doi.org/10.1590/S0074-02762008005000008>.
- Moody A. 2002. Rapid diagnostic tests for malaria parasites. *Clin. Microbiol. Rev.* 15:66–78. <http://dx.doi.org/10.1128/CMR.15.1.66-78.2002>.
- Cho CH, Nam MH, Kim JS, Han ET, Lee WJ, Oh JS, An SS, Lim CS. 2011. Genetic variability in *Plasmodium vivax* aldolase gene in Korean isolates and the sensitivity of the Binax Now malaria test. *Trop. Med. Int. Health* 16:223–226. <http://dx.doi.org/10.1111/j.1365-3156.2010.02691.x>.
- Ishengoma DS, Francis F, Mmbando BP, Lusingu JP, Magistrado P, Alifrangis M, Theander TG, Bygbjerg IC, Lemnge MM. 2011. Accuracy of malaria rapid diagnostic tests in community studies and their impact on treatment of malaria in an area with declining malaria burden in north-eastern Tanzania. *Malar. J.* 10:176. <http://dx.doi.org/10.1186/1475-2875-10-176>.
- Kim KH, Jang JW, Woo MK, Oh JS, Han ET, Lee WJ, An SS, Lim CS. 2011. Evaluation of four rapid diagnostic tests for the diagnosis of *Plasmodium vivax* in Korea. *Trop. Med. Int. Health* 16:1427–1431. <http://dx.doi.org/10.1111/j.1365-3156.2011.02846.x>.
- Larsson A, Sjoquist J. 1988. False-positive results in latex agglutination tests caused by rheumatoid factor. *Clin. Chem.* 34:767–768.
- World Health Organization. 2009. Malaria rapid diagnostic test performance: results of WHO product testing of malaria RDTs: round 1. World Health Organization, Geneva, Switzerland.
- World Health Organization. 2010. Malaria rapid diagnostic test performance: results of WHO product testing of malaria RDTs: round 2. World Health Organization, Geneva, Switzerland.

12. World Health Organization. 2010. Information note on interim selection criteria for procurement of malaria rapid diagnostic tests (RDTs). World Health Organization, Global Malaria Programme, Geneva, Switzerland.
13. Grobusch MP, Jelinek T, Hanscheid T. 1999. False positivity of rapid antigen detection tests for diagnosis of *Plasmodium falciparum* malaria: issue appears to be more complicated than presented. *J. Clin. Microbiol.* 37:3781–3782.
14. Iqbal J, Sher A, Rab A. 2000. *Plasmodium falciparum* histidine-rich protein 2-based immunocapture diagnostic assay for malaria: cross-reactivity with rheumatoid factors. *J. Clin. Microbiol.* 38:1184–1186.
15. Mishra B, Samantaray JC, Kumar A, Mirdha BR. 1999. Study of false positivity of two rapid antigen detection tests for diagnosis of *Plasmodium falciparum* malaria. *J. Clin. Microbiol.* 37:1233.
16. Peschken CA, Esdaile JM. 1999. Rheumatic diseases in North America's indigenous peoples. *Semin. Arthritis Rheum.* 28:368–391. [http://dx.doi.org/10.1016/S0049-0172\(99\)80003-1](http://dx.doi.org/10.1016/S0049-0172(99)80003-1).
17. Spector TD. 1990. Rheumatoid arthritis. *Rheum. Dis. Clin. North. Am.* 16:513–537.
18. Soltys AI, Axford JS, Sutton BJ. 1997. Rheumatoid factors: where are we now? *Ann. Rheum. Dis.* 56:285–286. <http://dx.doi.org/10.1136/ard.56.5.285>.