

# Evaluation of Matrix-Assisted Laser Desorption Ionization–Time of Flight Mass Spectrometry for Identification of Nontuberculous Mycobacteria from Clinical Isolates

Belén Rodríguez-Sánchez,<sup>a,b,c</sup> María Jesús Ruiz-Serrano,<sup>a,b,c</sup> Mercedes Marín,<sup>a,b,c,d</sup> Paula López Roa,<sup>a</sup> Marta Rodríguez-Créixems,<sup>a</sup> Emilio Bouza<sup>a,b,c,d</sup>

Clinical Microbiology and Infectious Diseases Department, Hospital General Universitario Gregorio Marañón, Madrid, Spain<sup>a</sup>; CIBER de Enfermedades Respiratorias (CIBERES CB06/06/0058), Madrid, Spain<sup>b</sup>; Instituto de Investigación Sanitaria Gregorio Marañón, Madrid, Spain<sup>c</sup>; Medicine Department, School of Medicine, Universidad Complutense de Madrid, Madrid, Spain<sup>d</sup>

**Matrix-assisted laser desorption ionization–time of flight mass spectrometry (MALDI-TOF MS) for the identification of nontuberculous mycobacterial (NTM) isolates was evaluated in this study. Overall, 125 NTM isolates were analyzed by MALDI-TOF and GenoType CM/AS. Identification by 16S rRNA/*hsp65* sequencing was considered the gold standard. Agreements between MALDI-TOF and GenoType CM/AS with the reference method were, respectively, 94.4% and 84.0%. In 17 cases (13.6%), results provided by GenoType and MALDI-TOF were discordant; however, the reference method agreed with MALDI-TOF in 16/17 cases (94.1%;  $P = 0.002$ ).**

Identification of nontuberculous mycobacteria (NTM) by conventional phenotypic tests requires long incubation periods that may take up to 12 weeks. The GenoType Mycobacterium CM/AS (Hain Lifescience GmbH, Nehren, Germany) is a test based on the amplification of a 23S rRNA gene region followed by reverse hybridization with specific probes that allows the identification of 40 of the most common NTM. However, the hybridization step requires high sequence homology, and even point mutations in the target regions may hamper the hybridization step, leading to unreliable results (1). Matrix-assisted laser desorption ionization–time of flight (MALDI-TOF) has been recently applied to the identification of a wide range of microorganisms (2), mainly clinically significant bacteria and fungi. Experience with MALDI-TOF for the rapid identification of NTM is very limited, mainly because of the number of identified species and because database information is scarce (3–5). In this study, a large collection of NTM clinical isolates and reference strains were analyzed using MALDI-TOF technology in comparison with the GenoType Mycobacterium CM/AS procedure and the reference procedure, 16S rRNA/*hsp65* gene sequencing.

From January 2010 to September 2014, all the NTM with purportedly clinical significance were grown on Lowenstein-Jensen agar and simultaneously identified by GenoType CM/AS assay (Hain Lifescience, GmbH, Nehren, Germany) and with MALDI-TOF MS technology (Bruker Daltonics, Bremen, Germany). Ten reference isolates were also included in this study as positive controls (Table 1). All the strains were also analyzed by 16S rRNA/*hsp65* sequencing as the reference method in order to resolve possible discrepancies.

**Genotypic identification.** Two gene regions were targeted for NTM identification: the 5' end of the 16S rRNA gene with the universal primers E8F (5'-AGAGTTTGATCCTGGCTCAG-3') and E533R (5'-TTACCGCGGCTGCTGGCA-3') (6) and a 439-bp fragment within the *hsp65* gene using the primers TB11 (5'-ACCAACGATGGTGTGTCAT) and TB12 (5'-CTTGTCGAACGCATACCT), as described by Telenti et al. (7). The amplification, purification, and sequencing process was carried out as previously described by Rodríguez-Sánchez et al. (8). Interpreta-

**TABLE 1** Identification of the reference strains by the two methods analyzed

| Mycobacterium ID by 16S rRNA/ <i>hsp65</i> | Reference no. | Concordance with:     |                       |
|--|---------------|-----------------------|-----------------------|
|  |               | MALDI-TOF ID          | GenoType CM/AS ID     |
| <i>M. avium</i>                            | NC103034      | <i>M. avium</i>       | <i>M. avium</i>       |
| <i>M. haemophilum</i>                      | DSM44634      | <i>M. haemophilum</i> | <i>M. haemophilum</i> |
| <i>M. malmoense</i>                        | DSM 44163     | <i>M. malmoense</i>   | <i>M. malmoense</i>   |
| <i>M. palustre</i>                         | DSM 44572     | <i>M. palustre</i>    | No ID                 |
| <i>M. shimoidei</i>                        | DSM 44152     | <i>M. shimoidei</i>   | <i>M. shimoidei</i>   |
| <i>M. simiae</i>                           | DSM 44165     | <i>M. simiae</i>      | <i>M. simiae</i>      |
| <i>M. smegmatis</i>                        | MC2 155       | <i>M. smegmatis</i>   | <i>M. smegmatis</i>   |
| <i>M. szulgai</i>                          | DSM 44166     | <i>M. szulgai</i>     | <i>M. szulgai</i>     |
| <i>M. triplex</i>                          | DSM 44626     | <i>M. triplex</i>     | No ID                 |
| <i>M. xenopi</i>                           | DSM 43995     | <i>M. xenopi</i>      | <i>M. xenopi</i>      |

tion of sequencing results was performed considering CLSI recommendations for the genus *Mycobacterium* (9). Results obtained by 16S rRNA/*hsp65* sequencing were considered the final microorganism identification.

**GenoType CM/AS identification.** The GenoType assay (Hain Lifescience, GmbH, Nehren, Germany) was applied to the NTM

Received 25 May 2015 Returned for modification 26 May 2015

Accepted 1 June 2015

Accepted manuscript posted online 10 June 2015

**Citation** Rodríguez-Sánchez B, Ruiz-Serrano MJ, Marín M, López Roa P, Rodríguez-Créixems M, Bouza E. 2015. Evaluation of matrix-assisted laser desorption ionization–time of flight mass spectrometry for identification of nontuberculous mycobacteria from clinical isolates. *J Clin Microbiol* 53:2737–2740. doi:10.1128/JCM.01380-15.

**Editor:** R. Patel

Address correspondence to Belén Rodríguez-Sánchez, mbelen.rodriguez@isgm.com.

Copyright © 2015, American Society for Microbiology. All Rights Reserved.

doi:10.1128/JCM.01380-15

TABLE 2 *Mycobacterium* isolates identified by 16S rRNA/hsp65 sequencing and their correlation with MALDI-TOF, grouped by score and GenoType results

| Mycobacterium species ID by 16S rRNA/hsp65 | No. (%) of isolates                                 |           |            |         |                                    |   |       |
|--|---|-----------|------------|---------|------------------------------------|---|-------|
|  | With MALDI-TOF ID at species level with a score of: |           |            |         | With discordant or no MALDI-TOF ID | With concordance with GenoType CM/AS ID | Total |
|  | ≥2.0  | ≥1.9      | ≥1.7       | <1.7    |                                    |   |       |
| <i>M. abscessus</i>                        | 7   | 7         | 9          | 0       | 0                                  | 7                                       | 9     |
| <i>M. arupense</i>                         | 0   | 1         | 1          | 0       | 0                                  | 0                                       | 1     |
| <i>M. avium</i>                            | 6   | 9         | 14         | 3       | 0                                  | 16                                      | 17    |
| <i>M. bohemicum</i>                        | 1   | 1         | 1          | 0       | 0                                  | 0                                       | 1     |
| <i>M. chelonae</i>                         | 5   | 5         | 6          | 0       | 0                                  | 6                                       | 6     |
| <i>M. fortuitum</i>                        | 10  | 14        | 17         | 0       | 0                                  | 16                                      | 17    |
| <i>M. gastri</i>                           | 1   | 1         | 1          | 0       | 0                                  | 1                                       | 1     |
| <i>M. goodii</i>                           | 4   | 8         | 13         | 2       | 3                                  | 17                                      | 18    |
| <i>M. haemophilum</i>                      | 0   | 1         | 1          | 0       | 0                                  | 1                                       | 1     |
| <i>M. intracellulare</i>                   | 2   | 3         | 4          | 1       | 0                                  | 4                                       | 5     |
| <i>M. kansasii</i>                         | 3   | 5         | 10         | 0       | 0                                  | 10                                      | 10    |
| <i>M. lentiflavum</i>                      | 6   | 8         | 8          | 0       | 0                                  | 7                                       | 8     |
| <i>M. mageritense</i>                      | 5   | 5         | 6          | 0       | 0                                  | 5                                       | 6     |
| <i>M. malmoense</i>                        | 1   | 1         | 1          | 0       | 0                                  | 1                                       | 1     |
| <i>M. marinum</i>                          | 1   | 1         | 1          | 0       | 0                                  | 1                                       | 1     |
| <i>M. massiliense</i>                      | 0   | 0         | 0          | 0       | 4                                  | 0                                       | 4     |
| <i>M. mucogenicum</i>                      | 2   | 2         | 2          | 0       | 0                                  | 2                                       | 2     |
| <i>M. palustre</i>                         | 0   | 0         | 1          | 0       | 0                                  | 0                                       | 1     |
| <i>M. peregrinum</i>                       | 3   | 3         | 3          | 0       | 0                                  | 1                                       | 3     |
| <i>M. porcinum</i>                         | 1   | 1         | 1          | 0       | 0                                  | 0                                       | 1     |
| <i>M. shimoidei</i>                        | 1   | 1         | 1          | 0       | 0                                  | 1                                       | 1     |
| <i>M. simiae</i>                           | 1   | 2         | 2          | 0       | 0                                  | 1                                       | 2     |
| <i>M. smegmatis</i>                        | 1   | 1         | 1          | 0       | 0                                  | 1                                       | 1     |
| <i>M. szulgai</i>                          | 1   | 1         | 1          | 0       | 0                                  | 1                                       | 1     |
| <i>M. terrae</i>                           | 1   | 1         | 1          | 0       | 0                                  | 1                                       | 1     |
| <i>M. triplex</i>                          | 0   | 1         | 1          | 0       | 0                                  | 0                                       | 1     |
| <i>M. xenopi</i>                           | 1   | 3         | 4          | 1       | 0                                  | 5                                       | 5     |
| Total                                      | 64 (51.2)   | 86 (68.8) | 111 (88.8) | 7 (5.6) | 7 (5.6)                            | 105 (84)                                | 125   |

TABLE 3 Resolution of discrepancies between MALDI-TOF and the GenoType CM/AS ID according to 16S rRNA/hsp65 ID<sup>a</sup>

| No. of isolates (n = 17) | MALDI-TOF ID               | GenoType CM/AS ID         | Mycobacterium ID by 16S rRNA/hsp65 |
|--------------------------|----------------------------|---------------------------|------------------------------------|
| 2                        | <i>M. abscessus</i>        | <i>M. chelonae</i>        | <i>M. abscessus</i>                |
| 1                        | <b><i>M. abscessus</i></b> | <b><i>M. chelonae</i></b> | <b><i>M. massiliense</i></b>       |
| 1                        | <i>M. arupense</i>         | <i>Mycobacterium</i> spp. | <i>M. arupense</i>                 |
| 1                        | <i>M. avium</i>            | <i>M. kansasii</i>        | <i>M. avium</i>                    |
| 1                        | <i>M. bohemicum</i>        | <i>M. scrofulaceum</i>    | <i>M. bohemicum</i>                |
| 1                        | <i>M. fortuitum</i>        | <i>Mycobacterium</i> spp. | <i>M. fortuitum</i>                |
| 1                        | <i>M. goodii</i>           | <i>M. szulgai</i>         | <i>M. goodii</i>                   |
| 1                        | <i>M. intracellulare</i>   | <i>M. avium</i> complex   | <i>M. intracellulare</i>           |
| 1                        | <i>M. lentiflavum</i>      | <i>Mycobacterium</i> spp. | <i>M. lentiflavum</i>              |
| 1                        | <i>M. mageritense</i>      | <i>M. fortuitum</i>       | <i>M. mageritense</i>              |
| 1                        | <i>M. palustre</i>         | <i>M. avium</i>           | <i>M. palustre</i>                 |
| 2                        | <i>M. peregrinum</i>       | <i>M. fortuitum</i>       | <i>M. peregrinum</i>               |
| 1                        | <i>M. porcinum</i>         | <i>M. fortuitum</i>       | <i>M. porcinum</i>                 |
| 1                        | <i>M. simiae</i>           | <i>M. goodii</i>          | <i>M. simiae</i>                   |
| 1                        | <i>M. triplex</i>          | <i>M. avium</i>           | <i>M. triplex</i>                  |

<sup>a</sup> The line in bold shows isolates with discordant ID by MALDI-TOF and the reference method.

isolates following the manufacturer's instructions. The test consists of PCR amplification of a 23S rRNA gene region, a reverse hybridization with specific probes immobilized on nitrocellulose strips, and detection of the banding patterns. The turnaround time for this technique is approximately 8 h.

**MALDI-TOF mass spectrometry.** Isolates were treated prior to MALDI-TOF analysis following the manufacturer's instructions. Under biosafety level 3 (BSL3) conditions, samples were inactivated for 30 min at 95°C, resuspended in 300 µl of deionized water plus 900 µl of ethanol, and centrifuged at 13,000 rpm, and the supernatants were discarded. Afterwards, the pellet was taken to a BSL2 laboratory in order to conclude the preprocessing steps (10). The MALDI-TOF settings used have already been described (11). All isolates were analyzed by MALDI-TOF MS, using a Microflex LT benchtop mass spectrometer (Bruker Daltonics, Bremen, Germany). FlexControl 3.3 and Maldi Biotyper 3.1 software (Bruker Daltonics, Bremen, Germany), respectively, were used for the control of the instrument and for spectrum analysis and comparison with the database (updated with 5,627 species plus the specific Mycobacteria Library v2.0 with 313 isolates from 127 species; Bruker Daltonics). When this method was performed, the total turnaround time until MALDI-TOF identification was obtained was 90 min.

Sensitivity and specificity values were compared using the McNemar test for paired samples, with two tails. Validity values were calculated with a 95% confidence interval (CI) following an exact binomial distribution. Data were analyzed using SPSS software package 18.0 (IBM, Chicago, IL, USA).

Identification of the 125 NTM isolates by MALDI-TOF and 16S rRNA/*hsp65* sequencing was compared, and agreement between methods at the species level was shown in 118/125 cases (94.4%) (Table 2). A score of >1.7 and/or consistent top four matches were considered reliable MALDI-TOF identification at the species level; none of the isolates analyzed was identified with this technique at the complex level. Discordant identifications (IDs) consisted of 4 minor errors (4 isolates of *Mycobacterium massiliense*, a species not represented in the database, identified as *Mycobacterium abscessus*). In addition, three *Mycobacterium gordonae* isolates could not be identified: two achieved no peaks and one yielded no reliable results.

Agreement between GenoType CM/AS and the reference method at the species level was shown in 105 cases (84.0%). Discrepancies were mainly due to the lack of identification of the less common *Mycobacterium* species (*Mycobacterium arupense*, *M. bohemicum*, *M. massiliense*, *M. palustre*, *M. porcinum*, and *M. triplex*) with GenoType CM/AS. Excluding the isolates from these species, agreement with the reference method was 90.5% (105/116) and 97.4% (113/116) for GenoType CM/AS and MALDI-TOF, respectively. This is actually one of the limitations of GenoType CM/AS, while the MALDI-TOF database is constantly expanding. In fact, the Mycobacteria Library v2.0 used in the present study includes 313 isolates from 127 species. To our knowledge, this is the first time this database has been evaluated.

MALDI-TOF MS results were also compared with those obtained by GenoType CM/AS. Both methods provided consistent identification in 108/125 cases (86.4%). In 16/17 of the remaining cases (94.1%;  $P = 0.002$ ) the gold standard was consistent with MALDI-TOF results (Table 3). Besides, the reference method provided an ID different than MALDI-TOF and the GenoType CM/AS in one case: a *M. massiliense* identified as *M. abscessus* by MALDI-TOF and as *M. chelonae* by GenoType CM/AS. According to Fangous et al. (12), the correct identification of *M. massiliense* can be performed with MALDI-TOF using peak analysis.

MALDI-TOF identification obtained a score  $\geq 2.0$  (valid at the species level) in 64 cases (51.2%). However, with a cutoff  $\geq 1.9$  a total of 68.8% of the isolates were correctly identified at the species level and 88.8% with a score between 1.9 and 1.7 (valid at the genus level). Moreover, 7 isolates (5.6%) with a score below 1.7 (not reliable according to the manufacturer) were identically identified by MALDI-TOF and 16S rRNA/*hsp65* sequencing at the species level (Table 2). Thus, a score  $\geq 1.7$  and/or consistent top four identifications can be considered a reliable NTM identification with MALDI-TOF without losing accuracy.

The analysis by species showed that *M. avium*, with 3 isolates identified with a score below 1.7, and *M. gordonae*, with 3 isolates not identified by MALDI-TOF and 2 isolates with a score below 1.7, were the most challenging NTM species for MALDI-TOF ID. We hypothesize that this could be due to the extraction method limitations or to the low number of isolates from these two species in the database. Both factors will require further analysis.

MALDI-TOF had already shown to be a reliable tool for the identification of mycobacterial isolates (3–5, 13, 14). However, both the improvement of the sample pretreatment and the update

TABLE 4 Identification of NTM isolates grouped by growth rate

| Mycobacterium group                        | MALDI-TOF ID at the species level with a score of: |                |                  |                | GenoType CM/AS ID at the species level |                |                  |                  | No. of isolates |
|--|--|----------------|------------------|----------------|--|----------------|------------------|------------------|-----------------|
|  | $\geq 2.0$   | $\geq 1.9$     | $\geq 1.7$       | Level          | $\geq 2.0$                             | $\geq 1.9$     | $\geq 1.7$       | Level            |                 |
| Rapid growers <sup>a</sup>                 | 68.0 (53.3–80.5)                                   | 100 (95.2–100) | 78.0 (64.0–88.5) | 100 (95.2–100) | 92.0 (80.7–97.8)                       | 100 (95.2–100) | 76 (61.8–86.9)   | 93.3 (85.1–97.8) | 50              |
| Slow growers, nonchromogens <sup>b</sup>   | 41.4 (23.5–61.0)                                   | 100 (96.2–100) | 62.1 (42.3–79.3) | 100 (96.2–100) | 86.2 (68.3–96.1)                       | 100 (96.2–100) | 86.2 (68.3–96.1) | 97.9 (92.7–99.7) | 29              |
| Slow growers, scotochromogens <sup>c</sup> | 39.4 (22.9–57.8)                                   | 100 (96.1–100) | 63.6 (45.1–79.6) | 100 (96.1–100) | 81.8 (64.5–93.0)                       | 100 (96.1–100) | 90.9 (75.7–98.1) | 97.8 (92.4–99.7) | 33              |
| Slow growers, photochromogens <sup>d</sup> | 38.5 (13.8–68.4)                                   | 100 (96.7–100) | 61.5 (31.6–86.1) | 100 (96.7–100) | 100 (75.2–100)                         | 100 (96.7–100) | 92.3 (64.0–99.8) | 99.1 (95.1–100)  | 13              |

<sup>a</sup> *M. abscessus*, *M. arupense*, *M. chelonae*, *M. fortuitum*, *M. mageritense*, *M. massiliense*, *M. mucogenicum*, *M. peregrinum*, *M. porcinum*, and *M. smegmatis*.  
<sup>b</sup> *M. avium*, *M. gastri*, *M. haemophilum*, *M. intracellulare*, *M. malmoense*, *M. pulchre*, *M. simonoi*, *M. terrae*, and *M. triplex*.  
<sup>c</sup> *M. bohemicum*, *M. gordonae*, *M. lentiflavum*, *M. szulgai*, and *M. xenopi*.  
<sup>d</sup> *M. kansasii*, *M. marinum*, and *M. simiae*.

of the specific mycobacterial database allowed the correct identification of 118 isolates from 26 different NTM species. The agreement with the reference method was 94.4% at the species level. Similar results have been reported, but the studies included fewer mycobacterial species (15, 16). Our results remain below the 97.2% reported by Balada-Llasat et al. (5), although their study included 13 different NTM species, half the number identified in this study.

In summary, MALDI-TOF identification of NTM species proved decisively superior to GenoType CM/AS both in accuracy (94.4% versus 84.0% according to the genomic sequencing) and in turnaround time (90 min versus 8 h). The high agreement of MALDI-TOF with 16S rRNA/*hsp65* sequencing allows its implementation in the microbiology laboratory for NTM identification, especially for rapidly growing NTM (Table 4), where MALDI-TOF has shown a very low rate of discrepant results (3 isolates [0.6%] in this study). MALDI-TOF is also cheaper than GenoType CM/AS (1.5€ versus 42€ per sample). Thus, our recommendation is to perform MALDI-TOF, when it is available, for all the samples suspected of being NTM and sequence only the isolates with unreliable or no MALDI-TOF ID, which may represent only 5.6% of isolates subjected to MALDI-TOF, according to our results.

#### ACKNOWLEDGMENTS

B.R.-S. is supported by the Miguel Servet Program (ISCIII-MICINN).

We thank Javier Rodriguez Granger from the Hospital Virgen de las Nieves (Granada, Spain), who kindly supplied the reference *Mycobacterium* strains, and Luis Alcalá for providing expert advice with the statistical analysis. We also thank Beckman-Coulter SLU staff and Francisco Soria Melguizo staff for providing technical and scientific support with Microflex LT MALDI-TOF during this study and Adrián Ruiz for his skillful processing of the *Mycobacterium* isolates.

#### REFERENCES

- Richter E, Rüscher-Gerdes S, Hillemann D. 2006. Evaluation of the GenoType *Mycobacterium* assay for identification of mycobacterial species from cultures. *J Clin Microbiol* 44:1769–1775. <http://dx.doi.org/10.1128/JCM.44.5.1769-1775.2006>.
- Clark AE, Kaleta EJ, Arora A, Wolk DM. 2013. Matrix-assisted laser desorption ionization-time of flight mass spectrometry: a fundamental shift in the routine practice of clinical microbiology. *Clin Microbiol Rev* 26:547–603. <http://dx.doi.org/10.1128/CMR.00072-12>.
- El Khéchine A, Couderc C, Flaudrops C, Raoult D, Drancourt M. 2011. Matrix-assisted laser desorption/ionization time-of-flight mass spectrometry identification of mycobacteria in routine clinical practice. *PLoS One* 6:e24720. <http://dx.doi.org/10.1371/journal.pone.0024720>.
- Saleeb PG, Drake SK, Murray PR, Zelazny AM. 2011. Identification of mycobacteria in solid-culture media by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol* 49:1790–1794. <http://dx.doi.org/10.1128/JCM.02135-10>.
- Balada-Llasat JM, Kamboj K, Pancholi P. 2013. Identification of mycobacteria from solid and liquid media by matrix-assisted laser desorption ionization-time of flight mass spectrometry in the clinical laboratory. *J Clin Microbiol* 51:2875–2879. <http://dx.doi.org/10.1128/JCM.00819-13>.
- Baker GC, Smith JJ, Cowan DA. 2003. Review and re-analysis of domain-specific 16S primers. *J Microbiol Methods* 55:541–555. <http://dx.doi.org/10.1016/j.mimet.2003.08.009>.
- Telenti A, Marchesi F, Balz M, Bally F, Böttger EC, Bodmer T. 1993. Rapid identification of mycobacteria to the species level by polymerase chain reaction and restriction enzyme analysis. *J Clin Microbiol* 31:175–178.
- Rodríguez-Sánchez B, Marín M, Sánchez-Carrillo C, Cercenado E, Ruiz A, Rodríguez-Crèixems M, Bouza E. 2014. Improvement of matrix-assisted laser desorption/ionization time-of-flight mass spectrometry identification of difficult-to-identify bacteria and its impact in the workflow of a clinical microbiology laboratory. *Diagn Microbiol Infect Dis* 79:1–6. <http://dx.doi.org/10.1016/j.diagmicrobio.2014.01.021>.
- CLSI. 2008. Interpretive criteria for identification of bacteria and fungi by DNA target sequencing; approved guideline. CLSI document MM18-A. Clinical and Laboratory Standards Institute, Wayne, PA.
- Mediavilla-Gradolph MC, De Toro-Peinado I, Bermúdez-Ruiz MP, García-Martínez MDL, Ortega-Torres M, Quezel-Guerraz NM, Palop-Borrás B. 2015. Use of MALDI-TOF MS for identification of nontuberculous *Mycobacterium* species isolated from clinical specimens. *BioMed Res Int* 2015:854078.
- Rodríguez-Sánchez B, Sánchez-Carrillo C, Ruiz A, Marín M, Cercenado E, Rodríguez-Crèixems M, Bouza E. 2014. Direct identification of pathogens from positive blood cultures using matrix-assisted laser desorption-ionization time-of-flight mass spectrometry. *Clin Microbiol Infect* 20:O421–O427. <http://dx.doi.org/10.1111/1469-0691.12455>.
- Fangous MS, Mougari F, Gouriou S, Calvez E, Raskine L, Cambau E, Payan C, Héry-Arnaud G. 2014. Classification algorithm for subspecies identification within the *Mycobacterium abscessus* species, based on matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol* 52:3362–3369. <http://dx.doi.org/10.1128/JCM.00788-14>.
- Pignone M, Greth KM, Cooper J, Emerson D, Tang J. 2006. Identification of mycobacteria by matrix-assisted laser desorption ionization-time-of-flight mass spectrometry. *J Clin Microbiol* 44:1963–1970. <http://dx.doi.org/10.1128/JCM.01959-05>.
- Lotz A, Ferroni A, Beretti JL, Dauphin B, Carbonnelle E, Guet-Revillet H, Veziris N, Heym B, Jarlier V, Gaillard JL, Pierre-Audigier C, Frapy E, Berche P, Nassif X, Bille E. 2010. Rapid identification of mycobacterial whole cells in solid and liquid culture media by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol* 48:4481–4486. <http://dx.doi.org/10.1128/JCM.01397-10>.
- Buchan BW, Riebe KM, Timke M, Kostrzewa M, Ledebauer NA. 2014. Comparison of MALDI-TOF MS with HPLC and nucleic acid sequencing for the identification of *Mycobacterium* species in cultures using solid medium and broth. *Am J Clin Pathol* 141:25–34. <http://dx.doi.org/10.1309/AJCPBPUBUDEW2OAG>.
- Mather CA, Rivera SF, Butler-Wu SM. 2014. Comparison of the Bruker Biotyper and Vitek MS matrix-assisted laser desorption ionization-time of flight mass spectrometry systems for identification of mycobacteria using simplified protein extraction protocols. *J Clin Microbiol* 52:130–138. <http://dx.doi.org/10.1128/JCM.01996-13>.





## AUTHOR CORRECTION

# Correction for Rodríguez-Sánchez et al., Evaluation of Matrix-Assisted Laser Desorption Ionization–Time of Flight Mass Spectrometry for Identification of Nontuberculous Mycobacteria from Clinical Isolates

Belén Rodríguez-Sánchez,<sup>a,b,c</sup> María Jesús Ruiz-Serrano,<sup>a,b,c</sup> Mercedes Marín,<sup>a,b,c,d</sup> Paula López Roa,<sup>a</sup> Marta Rodríguez-Créixems,<sup>a</sup> Emilio Bouza<sup>a,b,c,d</sup>

Clinical Microbiology and Infectious Diseases Department, Hospital General Universitario Gregorio Marañón, Madrid, Spain<sup>a</sup>; CIBER de Enfermedades Respiratorias (CIBERES CB06/06/0058), Madrid, Spain<sup>b</sup>; Instituto de Investigación Sanitaria Gregorio Marañón, Madrid, Spain<sup>c</sup>; Medicine Department, School of Medicine, Universidad Complutense de Madrid, Madrid, Spain<sup>d</sup>

Volume 53, no. 8, p. 2737–2740, 2015. Page 2740: The first paragraph of Acknowledgments should read “B.R.-S. is supported by the Miguel Servet Program (ISCIII-MICINN CP14/00220) from the Health Research Fund (FIS) of the Carlos III Health Institute (ISCIII), Madrid, Spain, partially financed by the European Regional Development Fund (FEDER) ‘A way of making Europe.’ The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.”

**Citation** Rodríguez-Sánchez B, Ruiz-Serrano MJ, Marín M, López Roa P, Rodríguez-Créixems M, Bouza E. 2016. Correction for Rodríguez-Sánchez et al., Evaluation of matrix-assisted laser desorption ionization–time of flight mass spectrometry for identification of nontuberculous mycobacteria from clinical isolates. *J Clin Microbiol* 54:1929. doi:10.1128/JCM.00815-16.

Copyright © 2016, American Society for Microbiology. All Rights Reserved.