

Comparison of Matrix-Based and Filter Paper-Based Systems for Transport of Plasma for HIV-1 RNA Quantification and Amplicon Preparation for Genotyping

Molly Levine,^a Ingrid Beck,^a Sheila Styrchak,^a Gregory Pepper,^b Lisa Frenkel^{a,b}

Center for Global Infectious Disease Research, Seattle Children's Research Institute, Seattle, Washington, USA^a; University of Washington, Seattle, Washington, USA^b

Two ambient-temperature, dry plasma transport systems, ViveST tubes and RNASound RNA sampling cards, and two extraction methods were compared to frozen plasma for HIV-1 RNA recovery. Significant RNA loss occurred: ViveST+MiniMag > ViveST+QIAamp > RNASound+QIAamp. RNA loss and low specimen volumes may affect the sensitivity of genotyping specimens with HIV-1 RNA of <4.70 log₁₀ copies/ml.

Plasma samples from HIV-1-infected individuals are monitored for drug resistance and HIV-1 RNA viral load (VL) before and during antiretroviral therapy (ART) in many treatment programs, generally at central laboratories (1). Quantification of HIV-1 RNA is generally performed on plasma samples frozen shortly after collection. However, the expense and logistics of maintaining this cold chain for transport from remote to central laboratories can be prohibitively expensive for low-resource communities. To more economically transport plasma, dried blood and plasma spots and matrix-based tubes have been developed to ship plasma at ambient temperatures. Dried blood spots prepared with whole-blood samples on Whatman 903 filter paper produce HIV drug resistance genotypes comparable to those with frozen plasma (2–6) and quantify plasma HIV-1 RNA loads with a mean loss of 0.23 log₁₀ copy (c)/ml (2). Given that Whatman 903 filter paper holds only 50 μl of plasma or whole blood per circle, methods able to transport larger specimen volumes at ambient temperatures and reliably preserve HIV-1 RNA might reduce costs and increase access to recommended monitoring of HIV-1 VL and genotypes.

The ViveST system (ViveBio, Alpharetta, GA) includes matrix-based tubes for ambient temperature storage and shipment of plasma specimens. These tubes can hold up to 1 ml of plasma, which is slowly loaded by dripping the plasma from a pipet onto the matrix. Once the plasma has completely dried (≥8 h), the tube containing the matrix and a desiccant is shipped at ambient temperatures. Storage of the plasma for 2 to 56 days in the ViveST tubes at ambient temperatures has been reported to decrease plasma HIV-1 RNA values by 0.23 to 0.53 log₁₀ c/ml compared to those for frozen plasma (7). In another study, plasma samples in ViveST tubes shipped to three U.S. sites at ambient temperatures lost HIV-1 RNA (mean, 1.07 log₁₀ c/ml) compared to frozen plasma samples (A. M. McClernon, G. Cloherty, and D. R. McClernon, presented at the Association for Molecular Pathology Annual Meeting, Washington, DC, 12 to 14 November 2014). Reproducible quantification of replicate aliquots from four plasma specimens led the latter authors to suggest that the loss in HIV-1 RNA is sufficiently consistent that a correction factor might provide reliable VL quantification. Variable preservation of HIV-1 RNA for resistance genotyping has been reported for plasma samples stored in ViveST tubes (7, 8). A study of 20 plasma samples, with a broad range of VL (3.49 to 5.41 log₁₀ c/ml), found

99.9% concordance of drug resistance mutations between frozen aliquots and corresponding aliquots dried in ViveST tubes (7). Another study of 50 plasma samples with relatively lower VL (median, 3.54 log₁₀ c/ml; interquartile range –IQR], 3.32 to 4.11) yielded genotypes from 36% of the specimens dried in ViveST tubes compared to 96% from frozen plasma, with 98.9% concordance of resistance genotypes between specimen storage conditions (8).

The RNASound blood RNA card system (FortiusBio, San Diego, CA) represent a filter paper-based system for storage and transportation of biological specimens. The cards hold 200 μl of whole blood or plasma, which after drying can be stored with a desiccant and shipped at ambient temperatures. To our knowledge, no studies of plasma stability in RNASound have been published. The manufacturer reports that viral RNA in dried plasma is stable for at least 7 days. In this study, we assessed HIV-1 RNA recovery by quantifying VL and amplifying HIV-1 *pol* from plasma aliquots of specimens stored in ViveST tubes (RNA extracted by two methods), stored on RNASound cards, and frozen.

To compare the ViveST and RNASound systems to frozen plasma, previously genotyped residual deidentified plasma samples from HIV-1-infected individuals were obtained from the University of Washington Clinical Virology Laboratory, Seattle, WA. Each frozen plasma sample was thawed and divided into three aliquots: (i) 140 μl underwent immediate extraction of RNA (QIAamp viral RNA minikit; Qiagen, Valencia, CA); (ii) a median of 950 μl (range, 400 to 1,000) was loaded onto a ViveST matrix tube and dried overnight in a biosafety hood; and (iii) 200 μl was loaded onto an RNASound card and dried for ≥2 h at ambient temperature.

Received 10 March 2016 Returned for modification 28 March 2016

Accepted 8 April 2016

Accepted manuscript posted online 13 April 2016

Citation Levine M, Beck I, Styrchak S, Pepper G, Frenkel L. 2016. Comparison of matrix-based and filter paper-based systems for transport of plasma for HIV-1 RNA quantification and amplicon preparation for genotyping. *J Clin Microbiol* 54:1899–1901. doi:10.1128/JCM.00541-16.

Editor: A. J. McAdam, Boston Children's Hospital

Address correspondence to Lisa Frenkel, lfrenkel@uw.edu.

Copyright © 2016, American Society for Microbiology. All Rights Reserved.




FIG 1 Comparison of HIV-1 RNA quantification and drug resistance genotyping following preservation of plasma samples in ViveST tubes (extracted by QIAamp or MiniMag) versus on RNASound cards (extracted by QIAamp) versus frozen (extracted by QIAamp). Aliquots of 35 clinical plasma specimens had RNA extracted from frozen plasma by QIAamp or were loaded into ViveST tubes, which following storage overnight at ambient temperatures had RNA extracted by QIAamp and MiniMag, or onto an RNASound card, which once dried had RNA extracted by QIAamp. The recovery/quantification of HIV-1 RNA with each storage and extraction method by a real-time PCR that amplified 114 bp of 5' long terminal repeat (LTR)-*gag* is shown. The HIV-1 RNA yield from each condition and volumes of plasma extracted is normalized to \log_{10} copies per milliliter of plasma. Amplification of 1,257 bp of HIV-1 *pol* for genotyping was performed using 10 μ l of extracted RNA, which included RNA from the following plasma volumes: frozen, 23 μ l; ViveST+QIAamp, 22 μ l; ViveST+MiniMag, 152 μ l; and RNASound+QIAamp, 33 μ l. Successful amplification of *pol* is indicated by open shapes and failed amplification by closed shapes.

To extract the RNA from the ViveST tube, each dried matrix was rehydrated with 1 ml of RNase-free water, and the plasma was recovered according to the manufacturer's instruction. The recovered plasma totaled \sim 940 μ l and was extracted immediately by two silica-based kits: (i) 140 μ l using a QIAamp kit and (ii) \sim 800 μ l using the NucliSENS MiniMag system (bioMérieux, Durham, NC), following the manufacturers' instructions. Each RNASound card was rehydrated with 720 μ l of Qiagen buffer AVL that includes carrier RNA plus 180 μ l of RNase-free water and extracted immediately following the instructions for the QIAamp kit. The RNA extracts (10 μ l) from each of the four methods were reverse transcribed separately (Thermo Fisher SuperScript III; Thermo Fisher Scientific, Waltham, MA) and the HIV-1 cDNA (20 μ l) was quantified in duplicate (10 μ l/reaction) by quantitative real-time PCR (qPCR) (9). The HIV-1 RNA yields from each condition and volumes of plasma extracted were normalized to \log_{10} copies per milliliter of plasma. Based on the volumes of plasma extracted, the lower limit of quantification of the real-time assay was 2.94 \log_{10} c/ml for ViveST+QIAamp and fresh extractions, 2.09 \log_{10} c/mL for ViveST+MiniMag, and 2.78 \log_{10} c/ml for RNASound+QIAamp.

The recovery of nucleic acids for HIV-1 drug resistance genotyping was compared separately by amplifying 10 μ l of each extracted RNA (derived from 23 μ l frozen plasma, a median of 22 μ l plasma from ViveST+QIAamp, a median of 152 μ l plasma from ViveST+MiniMag, and 33 μ l plasma from RNASound+QIAamp) in a nested PCR of HIV-1 *pol* (1,257 bp; 1st round forward primer CCTAGGAAAAAGGGCTGTTGGAAATGTGG and reverse primer AATCCCTGSATAAAATYTGACTTGCCARTT; 2nd round for-

ward primer GARAGACAGGCTAATTTTTTATGGGA and reverse primer AAYTTCTGTATATCATTGACAGTCCA) encoding amino acids in protease (PR) 1 through reverse transcriptase (RT) 251 using a TaKaRa PrimeScript One Step RT-PCR kit (version 2; TaKaRa, Shiga, Japan) and MyTaq DNA Polymerase (Biolone, London, United Kingdom). Amplification was assessed by electrophoresis in an agarose gel stained with ethidium bromide, and band intensity was assessed visually to determine if it appeared adequate for genotyping. HIV-1 genotyping was not attempted on these amplicons.

The 35 frozen plasma samples (HIV-1 subtypes B [33 samples], AG/G [1 sample], and A/AE [1 sample]) had a median HIV-1 RNA of 4.67 \log_{10} c/ml (range, 2.79 to 6.54). Drying the plasma diminished the HIV-1 RNA values compared to those for the frozen plasma, with median reductions of 1.3 \log_{10} c/ml (range, 0.5 to 3.2), 2.7 \log_{10} c/ml (range, 0.7 to 4.7), and 1.3 \log_{10} c/ml (range, 0.7 to 2.5) for ViveST+QIAamp, ViveST+MiniMag, and RNASound+QIAamp, respectively ($P < 0.001$ for each comparisons by F tests for overall significance) (Fig. 1). The loss of HIV-1 RNA observed from plasma samples stored in ViveST tubes was greater than previously reported (7, 10). The correlation of ViveST+QIAamp, ViveST+MiniMag, or RNASound+QIAamp to frozen plasma found R^2 values of 0.707, 0.063, and 0.703, respectively, which indicates that a conversion factor would provide moderate to poor confidence in providing an accurate VL. The ViveST+MiniMag extraction was particularly variable, and 9 (26%) samples failed to amplify by qPCR, with median plasma HIV-1 RNA of 4.64 \log_{10} c/ml (range, 3.91 to 6.09). As only one sample extracted by ViveST+QIAamp (HIV-1 RNA of 5.08 \log_{10} c/ml) and none of

the RNASound+QIAamp samples were below the limit of HIV-1 RNA qPCR detection, the MiniMag extraction appears inefficient for either capturing RNA or removing inhibitors of PCR.

Amplification of HIV-1 *pol* for genotyping was successful from 100% of the frozen plasma specimens and for 89% of ViveST+QIAamp, 71% of ViveST+MiniMag, and 83% of RNASound+QIAamp specimens (Fig. 1). This higher rate of amplification failure than that for qPCR is likely due to the reverse transcription/amplification of a longer RNA fragment (114 versus 1,257 bp) and/or reduced efficiency of the one-step RT-PCR used for genotyping compared to that of the two-step RT-PCR used for qPCR. The plasma samples failing amplification of *pol* correspond to those with lower VL ($P < 0.0001$ by Mann-Whitney U test), except specimen 9, in which, given the successful amplification of frozen plasma for genotyping, we suspect that inhibitors of PCR may have persisted following storage in ViveST and RNASound. Our results are consistent with those of previous studies that successfully genotyped 36% of samples after storage in ViveST tubes with RNA extracted using the NucliSENS EasyMag system, which has the same chemistry as the MiniMag system (8). ViveST tubes have the advantage of holding up to 1 ml of plasma, while RNASound cards hold 200 μ l. The larger volume appears advantageous for genotyping of samples with lower HIV-1 RNA levels by increasing the total number of HIV-1 RNA templates available for extraction, reverse transcription, and amplification. Given our HIV-1 RNA recovery, only a subset of the specimens would theoretically provide ≥ 50 HIV-1 cDNA templates/10 μ l of RNA extract (the minimum number of templates needed to detect minority mutant variants comprising $\leq 20\%$ of the viral population by consensus HIV genotyping): 54% of ViveST-Qiagen, 37% of ViveST-MiniMag, and 57% of RNASound card-Qiagen samples. Indeed, the VL was $< 4.70 \log_{10}$ c/ml in all of the RNASound+QIAamp and ViveST+QIAamp specimens, providing < 50 viral templates/10 μ l. In contrast, HIV-1 RNA recoveries from several specimens with $> 6 \log_{10}$ c/ml extracted from ViveST tubes by MiniMag were below the threshold, consistent with potential inhibitors of PCR.

Limitations of the study include the fact that the stability of the HIV-1 plasma RNA in the ViveST and RNASound systems was tested over a short duration of storage and only at $\sim 23^{\circ}\text{C}$ and $\sim 15\%$ humidity. Also, to save resources, amplified HIV-1 *pol* was not sequenced, so the detection of minority variants could not be compared.

In conclusion, both the ViveST tubes and RNASound cards preserved the majority of plasma HIV-1 RNA at ambient laboratory temperatures over a relatively short period of time. The proportion of HIV-1 RNA lost was moderately predictable with ViveST+QIAamp and RNASound+QIAamp and unpredictable with ViveST+MiniMag. The variable loss of HIV-1 RNA reduced the lower limit and reproducibility of HIV-1 RNA quantification. Furthermore, it diminished the input of viral templates for HIV-1 drug resistance genotyping, which reduces the likelihood that minority mutants would be detected by consensus sequencing when the plasma HIV-1 RNA is $< 4.70 \log_{10}$ c/ml. The variable loss of HIV-1 RNA using these storage and transport systems emphasizes the need for an improved inexpensive modality to transport plasma samples for virus quantification and genotyping.

ACKNOWLEDGMENTS

This work was supported by a generous gift of ViveST tubes by Richard Carroll at ViveBio, LLC, by the Clinical Research and Retrovirology Core of the Seattle Centers for AIDS Research (NIH P30 AI027757), by Seattle Children's Research Institute Strategic Funds, and by NIH funds from the International Maternal Pediatric Adolescent AIDS Clinical Trials Group (IMPAACT) Virology Developmental Laboratory Award (U01 AI068632).

The content is solely the responsibility of the authors and does not necessarily represent the official views of the NIH.

FUNDING INFORMATION

This work, including the efforts of Lisa M. Frenkel, was funded by HHS | National Institutes of Health (NIH) (P30 AI027757 and UM1 AI106716).

REFERENCES

1. Department of Health and Human Services. Guidelines for the use of antiretroviral agents in HIV-1-infected adults and adolescents. <https://aidsinfo.nih.gov/guidelines>.
2. Rottinghaus EK, Ugbena R, Diallo K, Bassey O, Azeez A, Devos J, Zhang G, Aberle-Grasse J, Nkengasong J, Yang C. 2012. Dried blood spot specimens are a suitable alternative sample type for HIV-1 viral load measurement and drug resistance genotyping in patients receiving first-line antiretroviral therapy. *Clin Infect Dis* 54:1187–1195. <http://dx.doi.org/10.1093/cid/cis015>.
3. Parry CM, Parkin N, Diallo K, Mwebaza S, Batamwita R, DeVos J, Bbosa N, Lyagoba F, Magambo B, Jordan MR, Downing R, Zhang G, Kaleebu P, Yang C, Bertagnolio S. 2014. Field study of dried blood spot specimens for HIV-1 drug resistance genotyping. *J Clin Microbiol* 52:2868–2875. <http://dx.doi.org/10.1128/JCM.00544-14>.
4. Ziemniak C, George-Agwu A, Moss WJ, Ray SC, Persaud D. 2006. A sensitive genotyping assay for detection of drug resistance mutations in reverse transcriptase of HIV-1 subtypes B and C in samples stored as dried blood spots or frozen RNA extracts. *J Virol Methods* 136:238–247. <http://dx.doi.org/10.1016/j.jviromet.2006.05.030>.
5. Masciotra S, Garrido C, Youngpairoj AS, McNulty A, Zahonero N, Corral A, Heneine W, de Mendoza C, Garcia-Lerma JG. 2007. High concordance between HIV-1 drug resistance genotypes generated from plasma and dried blood spots in antiretroviral-experienced patients. *AIDS* 21:2503–2511. <http://dx.doi.org/10.1097/QAD.0b013e3281c618db>.
6. Bertagnolio S, Soto-Ramirez L, Pilon R, Rodriguez R, Viveros M, Fuentes L, Harrigan PR, Mo T, Sutherland D, Sandstrom P. 2007. HIV-1 drug resistance surveillance using dried whole blood spots. *Antivir Ther* 12:107–113.
7. Lloyd RM, Jr, Burns DA, Huong JT, Mathis RL, Winters MA, Tanner M, De La Rosa A, Yen-Lieberman B, Armstrong W, Taega A, McCleron DR, Wetshtein JL, Friedrich BM, Ferguson MR, O'Brien W, Feorino PM, Holodniy M. 2009. Dried-plasma transport using a novel matrix and collection system for human immunodeficiency virus and hepatitis C virus virologic testing. *J Clin Microbiol* 47:1491–1496. <http://dx.doi.org/10.1128/JCM.02354-08>.
8. Diallo K, Lehotzky E, Zhang J, Zhou Z, de Rivera IL, Murillo WE, Nkengasong J, Sabatier J, Zhang G, Yang C. 2014. Evaluation of a dried blood and plasma collection device, SampleTanker, for HIV type 1 drug resistance genotyping in patients receiving antiretroviral therapy. *AIDS Res Hum Retroviruses* 30:67–73. <http://dx.doi.org/10.1089/aid.2013.0127>.
9. Arvold ND, Ngo-Giang-Huong N, McIntosh K, Suraseranivong V, Warachit B, Piyaworawong S, Changchit T, Lallemand M, Jourdain G, Perinatal HIV Prevention Trial (PHPT-1), Thailand. 2007. Maternal HIV-1 DNA load and mother-to-child transmission. *AIDS Patient Care STDS* 21:638–643. <http://dx.doi.org/10.1089/apc.2006.0169>.
10. Zandoni M, Cortes R, Diaz RS, Sucupira MC, Ferreira D, Inocencio LA, Villena C, Loveday C, Lloyd RM, Jr, Holodniy M. 2010. Comparative effectiveness of dried plasma HIV-1 viral load testing in Brazil using ViveST for sample collection. *J Clin Virol* 49:245–248. <http://dx.doi.org/10.1016/j.jcv.2010.08.017>.