


Multicenter Evaluation of Meridian Bioscience HSV 1&2 Molecular Assay for Detection of Herpes Simplex Virus 1 and 2 from Clinical Cutaneous and Mucocutaneous Specimens

Matthew L. Faron,^a Nathan A. Ledebner,^{a,b} Anami Patel,^c Safedin H. Beqa,^d Belinda Yen-Lieberman,^e Debra Kohn,^e Amy L. Leber,^f Donna Mayne,^g William I. Northern,^h  Blake W. Buchan^{a,b}

Medical College of Wisconsin, Milwaukee, Wisconsin, USA^a; Wisconsin Diagnostic Laboratories, Milwaukee, Wisconsin, USA^b; LeBonheur Children's Hospital, Memphis, Tennessee, USA^c; Pathology, Inc., Torrance, California, USA^d; Cleveland Clinic, Cleveland, Ohio, USA^e; Nationwide Children's Hospital, Columbus, Ohio, USA^f; Sacred Heart Hospital, Pensacola, Florida, USA^g; CompuNet Clinical Laboratories, Moraine, Ohio, USA^h

Herpes simplex virus (HSV) causes acute and relapsing symptoms characterized by ulcerative lesions. Laboratory diagnosis of HSV in cutaneous or mucocutaneous lesions has historically been performed with the use of viral cell culture systems; however, these tests are laborious and suffer decreased sensitivity for advanced-stage lesions. The recent availability of FDA-cleared moderately complex assays has resulted in the increased use of molecular diagnostics for the routine detection of HSV in superficial swab specimens. We performed a clinical evaluation of the recently FDA-cleared *illumigene* HSV 1&2 loop-mediated isothermal amplification (LAMP) assay (Meridian Bioscience, Cincinnati OH) for the detection and differentiation of HSV-1 and HSV-2 in cutaneous and mucocutaneous swab specimens. A total of 1,153 clinical swab specimens were collected and tested at 7 different clinical centers. Each specimen was tested for the presence of HSV-1 and HSV-2 using the *illumigene* assay, and results were compared to those of the enzyme-linked virus-inducible system (ELVIS) as the reference method. Overall, the *illumigene* assay demonstrated a sensitivity and specificity of 94.8% and 95.5%, respectively, for the detection of HSV-1. Detection of HSV-2 was similar, with a sensitivity of 98.9% and a specificity of 95.5%. Discrepant analysis was performed using an alternative molecular test (AmpliVue HSV1 + 2 assay; Quidel Molecular, San Diego, CA) on 91/99 specimens that were recorded as false positive (FP) or false negative (FN) compared to the reference method. In total, 57/78 (73%) FP and 9/13 (69%) FN *illumigene* results were supported by the AmpliVue result. The *illumigene* HSV 1&2 assay demonstrated high sensitivity and specificity to detect and differentiate HSV in clinical specimens and identified 57 additional specimens that were positive for HSV compared to culture. The use of LAMP eliminates the need for the cycling of temperatures and provides results in less than 60 min, with approximately 2 min of hands-on time per specimen.

Herpes simplex virus 1 (HSV-1) and 2 (HSV-2) are double-stranded DNA viruses of the *Herpesviridae* family and are a common cause of cutaneous and mucocutaneous oral and genital lesions (1, 2). Historically, HSV-1 has been associated with oral lesions while HSV-2 has more often been associated with genital lesions, although recently this epidemiology has been changing. Additionally, HSV can cause invasive disease, including sepsis and fulminate encephalitis in neonates that are infected by exposure to herpetic lesions present at the time of birthing (3). Infection with HSV-1 is common, with seroprevalence reaching 50% to 70% in developed countries and >80% in developing countries (4–6). The seroprevalence of HSV-2 varies from 5% to 40% (7, 8). While there is no cure for HSV infections, specific antiviral therapy exists to alleviate symptoms, shorten the duration of recurring outbreaks, and to treat life-threatening manifestations (9). Therefore, accurate laboratory diagnosis of these infections is important.

The diagnosis of HSV based solely on clinical exam is difficult. Typical HSV lesions can mimic those of other sexually transmitted infectious agents (e.g., *Treponema pallidum* and *Haemophilus ducreyi*), staphylococcal folliculitis, herpes zoster, and skin abrasion due to physical injury (10). Therefore, it is essential to confirm all possible herpetic infections with laboratory diagnostics. Classically, laboratory diagnosis of superficial HSV infections relied on the collection of vesicular exudate or scrapings from a lesion and inoculation into a susceptible cell line (Hep-2, A549, etc.). Follow-

ing incubation, cells are observed microscopically for viral-induced cytopathic effects (CPE) over 5 to 14 days (11). Standard cell culture was improved with the development of the enzyme-linked virus-inducible system (ELVIS; Diagnostic Hybrids, Athens, OH), which increased sensitivity and reduced turnaround time (TAT) to approximately 24 h (12, 13). Overall, the sensitivity and specificity of ELVIS has been reported to range from 88% to 100% and 98% to 100%, respectively (14, 15). Importantly, the sensitivity of all culture-based diagnostics is dependent upon the presence and load of viable virus in a specimen, which is affected by the age of the lesion and specimen handling (16–18).

Molecular assays include laboratory-developed and FDA-cleared multiplex assays for simultaneous detection of and differ-

Received 4 March 2016 Returned for modification 2 April 2016

Accepted 13 May 2016

Accepted manuscript posted online 18 May 2016

Citation Faron ML, Ledebner NA, Patel A, Beqa SH, Yen-Lieberman B, Kohn D, Leber AL, Mayne D, Northern WI, Buchan BW. 2016. Multicenter evaluation of Meridian Bioscience HSV 1&2 molecular assay for detection of herpes simplex virus 1 and 2 from clinical cutaneous and mucocutaneous specimens. *J Clin Microbiol* 54:2008–2013. doi:10.1128/JCM.00483-16.

Editor: A. M. Caliendo, Rhode Island Hospital

Address correspondence to Blake W. Buchan, bbuchan@mcw.edu.

Copyright © 2016, American Society for Microbiology. All Rights Reserved.

entiation of HSV-1 and HSV-2 in superficial specimens (19–21). These assays are more sensitive and reduce TAT compared to viral culture and ELVIS (22–24). We present the first clinical evaluation, to our knowledge, of the recently FDA-cleared *illumigene* HSV 1&2 assay (Meridian Bioscience, Cincinnati, OH). In this study, a total of 1,158 residual cutaneous or mucocutaneous swab specimens from 7 clinical sites were collected for HSV testing. All specimens were tested using the *illumigene* HSV 1&2 assay, and results were compared to ELVIS as the reference method. Overall performance was determined by calculating the sensitivity and specificity of the assay. A subset of specimens was also analyzed in parallel using a real-time PCR (RT-PCR)-based laboratory-developed test (LDT) as a molecular comparator.

MATERIALS AND METHODS

Collection of HSV specimens. Residual swab specimens collected in viral transport media (including M4, M4RT, M6, UniTranz-RT, UTM, universal viral transport [UVT], and viral transport media [VTM]) and submitted to the clinical laboratory for routine HSV testing were enrolled in the study. Swabs from lesions of cutaneous (arms/hands, breast, back, abdomen, penile) and mucocutaneous (anorectal, vaginal, cervical, nasal, ocular, urethral, and oral) sources were acceptable specimens; no cerebral spinal fluid specimens were enrolled. All *illumigene* testing was performed on fresh specimens at each enrollment site. Corresponding ELVIS testing was performed on fresh specimens when possible; however, logistical aspects of the study design resulted in 589 specimens (50.8%) being tested by ELVIS following a freeze cycle. Fresh specimens were defined as specimens stored at 2°C to 4°C for up to 72 h. Frozen specimens were stored at –80°C before testing and were only thawed once. Fresh and frozen testing modalities are acceptable according to manufacturer recommendations. Specimen collection and testing were conducted at 7 different geographical locations (Sacred Heart Hospital, Pensacola, FL; The Medical College of Wisconsin, Milwaukee, WI; Pathology Inc., Torrance, CA; Cleveland Clinic, Cleveland, OH; CompuNet Clinical Laboratories, Moraine, OH; LeBonheur Children's Hospital, Memphis, TN; and Nationwide Children's Hospital, Columbus, OH). All testing for this study was performed using deidentified residual specimens after standard of care testing was completed. Each site performed testing in accordance with site-specific institutional review board approved protocols.

Enzyme-linked virus-inducible system. ELVIS testing was performed using the ELVIS D³ typing kit (Diagnostic Hybrids, Athens, OH) in accordance with the package insert protocol using standard shell vials (Diagnostic Hybrids, Athens, OH). The presence of blue cells due to the ELVIS β-galactosidase reaction was used as an indication of HSV-positive specimens. Positive specimens were then typed using fluorescently tagged antibodies contained within the kit. Positive specimens that could not be typed were reported as untypeable. Sites C, E, F, and G were unable to perform ELVIS testing onsite, so specimens were shipped frozen to site A for reference testing.

***illumigene* HSV 1&2 assay.** The *illumigene* HSV 1&2 assay is a molecular assay that utilizes loop-mediated isothermal DNA amplification (LAMP). The assay targets a 208-bp sequence of the HSV-1 glycoprotein G and a 189-bp sequence of the HSV-2 glycoprotein G gene. Two test devices are used for the assay: one containing lyophilized reagents for detection of HSV-1 and the second containing reagents for detection of HSV-2. Testing using *illumigene* was performed following manufacturer recommendations. Specimens are reported as positive for HSV-1 or HSV-2 target DNA, negative, or invalid. All invalid results were repeated once by repeating preparation from the patient specimen.

Discrepant resolution. Specimens that demonstrated discrepant results between *illumigene* HSV 1&2 and ELVIS were resolved using the FDA-cleared AmpliVue HSV1+2 assay (Quidel Molecular, San Diego, CA) as an alternative molecular test. All discrepant testing was performed

TABLE 1 HSV prevalence in cutaneous specimens by patient age

Age (yr)	<i>illumigene</i> HSV 1			<i>illumigene</i> HSV 2		
	Total no.	Total no. positive ^a	Prevalence (%) ^b	Total no.	Total no. positive ^a	Prevalence (%) ^b
≤5	38	12	31.6	38	1	2.6
6–18	42	18	42.9	42	4	9.5
19–55	177	20	11.3	177	37	20.9
>55	49	4	8.2	49	13	26.5
Not provided	0	0	0.0	0	0	0.0

^a Positivity was calculated based on the *illumigene* results.

^b Specimens resulting in invalid results were removed for prevalence analysis (no.).

blinded at a single reference site (Meridian Bioscience Inc.), and all tests were performed from frozen specimens.

Laboratory-developed real-time PCR assay. A laboratory-developed test based on HSV-1 and HSV-2 analyte-specific reagents (Cepheid, Sunnyvale, CA) and SmartMix HM beads (Cepheid, Sunnyvale, CA) that contained *Taq* polymerase, deoxynucleoside triphosphate (dNTPs), and MgCl₂ (mastermix) was used as a molecular comparator for a subset of fresh clinical specimens ($n = 363$). Specimens in VTM were prepared by 10 min of heat inactivation at 95°C, and 5 μl of heat-treated specimen was combined with 15 μl of mastermix. Amplification and detection were carried out on a SmartCycler II (Cepheid, Sunnyvale, CA) with the following cycling conditions: 1 cycle heating at 95°C for 15 s and 50 cycles of 95, 57, and 72°C for 15 s each. Specimens were identified as positive if either fluorescence for HSV-1 or HSV-2 targets had a cycle threshold (C_T) of ≤45 and both the positive and negative external controls passed.

Statistical analysis. Results from the *illumigene* HSV 1&2 assay were compared to those of ELVIS culture. Performance characteristics, including sensitivity and specificity, were calculated using standard methods. Ninety-five percent confidence intervals (CIs) were calculated by using a binomial expansion.

RESULTS

Specimen characteristics and prevalence. A total of 1,158 specimens that met study criteria were used to determine assay performance. Overall, the prevalences of HSV-1 and HSV-2 based on ELVIS results were 21.7% and 15.5%, respectively. Enrolled specimens were collected from pediatric (≤18 years of age, $n = 208$) and adult (>18 years of age, $n = 947$) patients. Patient age was unknown for three specimens. The prevalences of HSV-1 and HSV-2 varied between age groups, lesion types, and virus types, ranging from 0% to 43% (Tables 1 and 2). Two specimens were positive for both HSV-1 and HSV-2 by *illumigene* and were removed from statistical calculations due to an inability to detect coinfection using ELVIS. One specimen that was positive for HSV by ELVIS but could not be typed was removed from analysis. The final number of specimens used to establish the performance of *illumigene* HSV 1&2 was 1,153.

Comparison of *illumigene* to ELVIS for the identification of HSV-1 and HSV-2. The performance of the *illumigene* assay to detect HSV-2 was calculated by comparing the *illumigene* assay result to that of ELVIS, which was reported as the true value (Table 3). The *illumigene* reported a positive result for HSV-2 in 179/181 (98.9%) specimens classified as HSV-2 by ELVIS. Target nucleic acid was not detected in the remaining 2 specimens (i.e., false negative [FN]). The *illumigene* assay identified an additional 44 specimens as positive for HSV-2, which were classified as negative by ELVIS (i.e., false positive [FP]). These data resulted in an over-

TABLE 2 HSV prevalence in mucocutaneous specimens by patient age

Age (yr)	<i>illumigene</i> HSV 1			<i>illumigene</i> HSV 2		
	Total no.	Total no. positive ^a	Prevalence (%) ^b	Total no.	Total no. positive ^a	Prevalence (%) ^b
≤5	47	8	17.0	47	0	0.0
6 to 18	81	17	21.0	81	15	18.5
19 to 55	627 (1)	137	21.8	628	138	22.0
>55	91 (2)	21	23.1	92 (1)	16	17.4
Not provided	3	0	0.0	2 (1)	0	0.0

^a Positivity was calculated based on the *illumigene* results.

^b Specimens resulting in invalid results were removed for prevalence analysis (no.).

all sensitivity of 98.9% (CI, 96% to 100%) and specificity of 95.5% (CI, 94% to 97%) for HSV-2.

ELVIS typing uses the same fluorophore for typing HSV-1 and HSV-2; therefore, a specimen that is positive for HSV-2 is not tested for HSV-1. Given this limitation, all specimens typed as HSV-2 by ELVIS were removed from the data analysis used to calculate the performance of the *illumigene* HSV 1&2 assay for HSV-1 detection. Removal of the 181 specimens typed as HSV-2 by ELVIS decreased the total specimens included for HSV-1 to 974. The *illumigene* assay reported a positive result for HSV-1 in 200/211 (94.8%) specimens also typed as HSV-1 by ELVIS. The *illumigene* detected HSV-1 in an additional 34 specimens that were classified as negative by ELVIS. These data resulted in a sensitivity of 94.8% (95% CI, 91% to 97%) and a specificity of 95.5% (95% CI, 94% to 97%) for the detection of HSV-1 compared with ELVIS (Table 3).

The sensitivity and specificity of *illumigene* for HSV were statistically equivalent between the pediatric (<18 years of age) and adult groups ($P > 0.05$), with the exception of specificity for HSV-2. In adults, the specificity was determined to be 94.8% (740/781) while in pediatrics the specificity was 98.4% (188/191) ($P = 0.045$). Following discrepant resolution, no statistical difference in assay performance was observed between age groups. Therefore,

the initial difference may be attributable to the reduced sensitivity of the ELVIS reference method in adults.

Culture-based prevalences of HSV-1 (7% to 28%) and HSV-2 (6% to 30%) were variable among the clinical trial sites. This may have impacted site-specific performance statistics; specifically, one of two sites with <90% sensitivity for HSV-1 (site C) enrolled only 90 specimens. The other site with <90% sensitivity for HSV-1 (site E) enrolled 254 specimens but was also the only site to demonstrate <100% sensitivity for detection of HSV-2. These data would suggest a potential site-specific difference in laboratory practice or specimen collection that negatively impacted the assay results.

Performance of *illumigene* HSV 1&2 for lesion source. The performances of molecular assays can be affected by inhibitory substances present in clinical specimens. Eight types of specimens were enrolled in this study that represented various anatomic locations and matrices, with the most common being genital, including cutaneous (penile) and mucocutaneous (vaginal/cervical), followed by superficial skin lesions and oral sources (Table 4).

The *illumigene* assay performed similarly independent of specimen source or location. Both genital cutaneous and mucocutaneous specimens demonstrated >99% sensitivity for HSV-2 detection. Sensitivity for the detection of HSV-1 was higher for

TABLE 3 Performance of *illumigene* HSV 1&2 assay compared to that of ELVIS culture

Analyte	Site	No. TP ^a	No. TN ^b	No. FP	No. FN	Total no.	Sensitivity (%) (95% CI)	Specificity (%) (95% CI)
HSV-1	A	9	38	0	0	47	100 (63–100)	100 (88–100)
	B	64	217	11	1	293	98.5 (91–100)	95.2 (91–97)
	C	9	61	2	2	74	81.8 (48–97)	96.8 (88–99)
	D	52	147	6	0	205	100 (91–100)	96.1 (91–98)
	E	34	173	14	7	228	82.9 (67–92)	92.5 (88–96)
	F	1	14	0	0	15	100 (5–100)	100 (73–100)
	G	31	79	1	1	112	96.9 (82–100)	98.8 (92–100)
	Total	200	729	34	11	974 ^c	94.8 (91–97)	95.5 (94–97)
HSV-2	A	20	45	2	0	67	100 (80–100)	95.7 (84–99)
	B	70	281	12	0	363	100 (94–100)	95.9 (93–98)
	C	16	68	6	0	90	100 (76–100)	91.9 (82–97)
	D	38	199	6	0	243	100 (88–100)	97.1 (93–99)
	E	24	214	14	2	254	92.3 (73–99)	93.8 (90–96)
	F	2	15	0	0	17	100 (20–100)	100 (75–100)
	G	7	108	4	0	119	100 (56–100)	96.4 (90–99)
	Total	177	930	44	2	1,153 ^d	98.9 (96–100)	95.5 (94–97)

^a TP, true positive.

^b TN, true negative.

^c HSV-2 ELVIS-positive specimens were removed from HSV-1 analysis, as positive HSV-2 cells are not tested for HSV-1 coinfection following the ELVIS protocol.

^d Two specimens were invalid after a repeat test.

TABLE 4 Performance of *illumigene* HSV 1&2 assay based on site of lesion

Location	Analyte	No. TP	No. TN	No. FP	No. FN	Total	Sensitivity (%) (95% CI)	Specificity (%) (95% CI)
Genital ^a (cutaneous)	HSV-1	6	63	1	2	72	75.0 (41–93)	98.4 (92–100)
	HSV-2	20	64	8	0	92	100 (84–100)	88.9 (80–94)
Skin lesion	HSV-1	42	144	5	1	192	97.7 (88–100)	96.6 (92–99)
	HSV-2	22	187	5	0	214	100 (85–100)	97.4 (94–99)
Genital ^b (mucocutaneous)	HSV-1	93	376	17	7	493	93.0 (86–97)	95.7 (93–97)
	HSV-2	128	467	27	1	623	99.2 (96–100)	94.5 (92–96)
Oral	HSV-1	44	81	9	0	134	100 (92–100)	90.0 (82–95)
	HSV-2	0	132	2	1	135	0.0 (0–79)	98.5 (95–100)
Anorectal	HSV-1	7	33	0	0	40	100 (65–100)	100 (90–100)
	HSV-2	7	37	2	0	46	100 (65–100)	94.9 (83–99)
Nasal	HSV-1	8	9	1	0	18	100 (68–100)	90.0 (60–98)
	HSV-2	0	18	0	0	18	NA ^c	100 (82–100)
Ocular	HSV-1	0	19	0	1	20	0.0 (0–79)	100 (83–100)
	HSV-2	0	20	0	0	20	NA	100 (84–100)
Urethral	HSV-1	0	4	1	0	5	NA	80.0 (38–96)
	HSV-2	0	5	0	0	5	NA	100 (57–100)

^a Penile.^b Vaginal/cervical.^c NA, not available.

mucocutaneous specimens (93.0%) than for cutaneous specimens (75.0%); however, this may be the result of a comparatively low cutaneous genital HSV-1 prevalence, with only 8/72 specimens positive for HSV-1. The sensitivity for HSV-1 was 100% in oral lesions. Detection from superficial skin lesions was also sensitive for HSV-1 (97.7%) and HSV-2 (100%).

Overall, the specificity from each source was consistent with the majority of sources having >90% specificity prior to discrepant resolution. Two sources, cutaneous genital (HSV-2) and urethral (HSV-1) were below 90% specificity. In each of these data sets, prevalence was low. Collections of specimens from anorectal, nasal, ocular, and urethral sources were included in the data, but total specimen enrollment from these sources was low (5 to 46 specimens), which may not rigorously test the assay's performance for these sites.

Discrepant analysis. Discrepant analysis was performed on all specimens with discordant results between *illumigene* and ELVIS using the AmpliVue HSV1+2 assay (Quidel, San Diego, CA). The AmpliVue results were in agreement with *illumigene* in 78% (25/32) of HSV-1 specimens and in 82.0% (32/39) of HSV-2 specimens that were initially classified as FP (Table 5). Additionally, AmpliVue results were in agreement with *illumigene* in 80% (8/10) of HSV-1 specimens and in 50% (1/2) of HSV-2 specimens that were initially classified as FN. These data suggest that the majority of FP results likely contained the HSV amplicon, which is consistent with the increased sensitivity of molecular assays compared to that of cell culture (22, 25).

Thresholds of detection for ELVIS and the *illumigene* assay for the detection of HSV. A subset of specimens was also tested using a laboratory-developed RT-PCR test, which is used as the standard of care at site B (see Materials and Methods). Cycle threshold (C_T) values were collected for all fresh specimens, and the percent positive agreements (PPAs) for ELVIS and *illumigene* were calculated at three C_T categories: ≤ 30 , ≤ 35 , and ≤ 45 (Table 6). Frozen specimens were not included in this analysis since a freeze-thaw cycle may reduce the sensitivity of ELVIS and prevent a fair comparison of the assays.

Among the 254 fresh specimens tested, 119 were identified as positive for HSV by the LDT. Of the 119 specimens that tested positive by LDT, 67.2% (80/119) had C_T scores of ≤ 30 , indicating a relatively high viral burden. The ELVIS results were in agreement with the LDT results for 78/80 (97.5%) specimens in this group, and the *illumigene* assay agreed with the LDT for all 80 (100%). An additional 22 specimens were positive according to the LDT with a C_T of 30 to 35, indicating a medium to low viral burden. Including these specimens in the analysis, ELVIS detected HSV in 91/102 (89.2%) specimens, while *illumigene* detected HSV in 99/102 (97.0%) specimens with C_T values of ≤ 35 . Finally, 17 specimens were positive by LDT with C_T values of >35 to 45, indicating a very low viral burden. The *illumigene* assay detected HSV in 5/17 (29.4%) specimens, while ELVIS detected HSV in only 2/17 (11.8%). The two positive ELVIS cultures were also positive by *illumigene*. Including all of the LDT-positive HSV specimens, ELVIS detected 93/119 (78.2%) specimens whereas *illumigene* detected 104/119 (87.4%) specimens.

TABLE 5 Discrepant analysis using Quidel AmpliVue HSV1+2 assay as an alternative molecular test

Analyte	No. of specimens	<i>Illumigene</i> result	ELVIS result ^a	AmpliVue result
HSV-1 ^b	25	Positive	Negative	Positive
	7	Positive	Negative	Negative
	2	Negative	Positive	Positive
	8	Negative	Positive	Negative
HSV-2 ^c	32	Positive	Negative	Positive
	7	Positive	Negative	Negative
	1	Negative	Positive	Positive
	1	Negative	Positive	Negative

^a Five ELVIS results were typed HSV-1, but both molecular assays reported positive for HSV-2.^b Three of 45 HSV-1 discrepant specimens were not performed on the AmpliVue HSV1+2 assay.^c Five of 46 HSV-2 discrepant specimens were not performed on the AmpliVue HSV1+2 assay.

TABLE 6 Percent positive agreement of ELVIS and *illumigene* compared to LDT based on RT-PCR cycle threshold value

Assay	PPA (no.) of specimens with C_T value ^a		
	$\leq 30 C_T$ (n = 80)	$\leq 35 C_T$ (n = 102)	$\leq 45 C_T$ (All positives, n = 119)
ELVIS	97.5 (78)	89.2 (91)	78.2 (93)
<i>illumigene</i>	100 (80)	97.0 (99)	87.4 (104)

^a Only specimens tested from fresh media were included in this analysis.

DISCUSSION

The data from our study are consistent with other publications that demonstrate the increased sensitivity of molecular detection methods for HSV compared to that of cell culture. This is illustrated by the large number of specimens testing positive by *illumigene* but negative by ELVIS. To support the presence of HSV nucleic acid in this cohort, 83 specimens with sufficient volume for additional analysis were tested using another isothermal amplification assay based on a different HSV genetic target (AmpliVue HSV1+2; Quidel). The AmpliVue result was in agreement with *illumigene* for 57 of 83 (78.1%) specimens, generating initial discrepant results between *illumigene* and ELVIS. The remaining eight specimens were unavailable for discrepant analysis (7 mucocutaneous and 1 cutaneous). A limitation of the discrepant analysis was that all specimens were frozen prior to AmpliVue testing. This may have reduced the sensitivity of AmpliVue since freeze-thaw cycles may reduce sensitivity of molecular assays. Despite this limitation, the overall specificity of *illumigene* post discrepant resolution rose to 98.8% for HSV-1 and HSV-2.

The impact of viral burden present in the clinical specimen on the sensitivity of culture was clearly demonstrated by the PPA when comparing ELVIS to the LDT or *illumigene* in relationship to C_T values. ELVIS culture detected HSV in only 59.1% of specimens with high C_T values (30 to 35) and was further diminished when C_T values were ≥ 35 . The loss of low-level detection by ELVIS may be clinically significant since viral burden can differ based on several factors, such as age of the lesion, state of infection (primary versus reactivation), virus (HSV-2 versus HSV-1), and quality of specimen collection (17, 26, 27). If ELVIS was used as the standard of care at site B, 26 patients would have been misdiagnosed, potentially leading to undertreatment and an increase in the spread of the virus. Importantly, the *illumigene* assay was positive in 11 of these ELVIS negative specimens. Without clinical and patient data, such as stage of lesion and history of antiviral therapy, it is difficult to assess the relevance of these very low-viral-load specimens. These may represent nonviable virus in resolving lesions, environmental (amplicon) contamination, or poorly collected specimens.

There were some limitations to this study. The FDA claim for the *illumigene* assay includes a wide variety of cutaneous and mucocutaneous sources, including skin, genital, anorectal, ocular, urethral, nasal, and oral lesions. During this evaluation, a limited number of nasal, ocular, and urethral specimens were received and tested. Culture-based prevalence of HSV for all three sources was 0%, preventing an accurate calculation of *illumigene* sensitivity in these specimens. Similarly, only a single oral specimen was positive for HSV-1 by ELVIS. In these instances, the calculated sensitivities, specificities, and 95% confidence intervals may not be a true reflection of *illumigene* performance.

In addition to the increased sensitivity offered by *illumigene*, the hands-on time and total time to result are also reduced. ELVIS and other cell culture methods require manual setup and microscopic examination, which can be subjective and requires a minimum of 24 h for final results. The *illumigene* assay provides a result within 1 h from the time the specimen is received by the laboratory, requires approximately 2 min of hands-on time per specimen, and is capable of simultaneous analysis of 5 patient specimens per *illumipro-10* analyzer. This test may be a good fit for laboratories that want the flexibility of on-demand or small-batch testing with a “moderate complexity” molecular assay.

FUNDING INFORMATION

This work, including the efforts of Blake Wade Buchan, Matthew Faron, Nathan A. Ledeboer, Anami Patel, Safedin Beqa, Belinda Yen-Lieberman, Debra Kohn, Amy Leber, Donna Mayne, and William Northern, was funded by Meridian Biosciences Inc.

REFERENCES

- Garland SM, Steben M. 2014. Genital herpes. *Best Pract Res Clin Obstet Gynaecol* 28:1098–1110. <http://dx.doi.org/10.1016/j.bpobgyn.2014.07.015>.
- Whitley RJ, Roizman B. 2001. Herpes simplex virus infections. *Lancet* 357:1513–1518. [http://dx.doi.org/10.1016/S0140-6736\(00\)04638-9](http://dx.doi.org/10.1016/S0140-6736(00)04638-9).
- Pinninti SG, Kimberlin DW. 2014. Preventing herpes simplex virus in the newborn. *Clin Perinatol* 41:945–955. <http://dx.doi.org/10.1016/j.clp.2014.08.012>.
- Gupta R, Warren T, Wald A. 2007. Genital herpes. *Lancet* 370:2127–2137. [http://dx.doi.org/10.1016/S0140-6736\(07\)61908-4](http://dx.doi.org/10.1016/S0140-6736(07)61908-4).
- Nahmias AJ, Lee FK, Beckman-Nahmias S. 1990. Sero-epidemiological and -sociological patterns of herpes simplex virus infection in the world. *Scand J Infect Dis Suppl* 69:19–36.
- Xu F, Sternberg MR, Kottiri BJ, McQuillan GM, Lee FK, Nahmias AJ, Berman SM, Markowitz LE. 2006. Trends in herpes simplex virus type 1 and type 2 seroprevalence in the United States. *JAMA* 296:964–973. <http://dx.doi.org/10.1001/jama.296.8.964>.
- LeGoff J, Pere H, Belec L. 2014. Diagnosis of genital herpes simplex virus infection in the clinical laboratory. *Virology* 11:83. <http://dx.doi.org/10.1186/1743-422X-11-83>.
- Schulte JM, Bellamy AR, Hook EW, III, Bernstein DI, Levin MJ, Leone PA, Sokol-Anderson ML, Ewell MG, Wolff PA, Heineman TC, Belshe RB. 2014. HSV-1 and HSV-2 seroprevalence in the United States among asymptomatic women unaware of any herpes simplex virus infection (Herpevac trial for women). *South Med J* 107:79–84. <http://dx.doi.org/10.1097/SMJ.0000000000000062>.
- James SH, Prichard MN. 2014. Current and future therapies for herpes simplex virus infections: mechanism of action and drug resistance. *Curr Opin Virol* 8:54–61. <http://dx.doi.org/10.1016/j.coviro.2014.06.003>.
- DiCarlo RP, Martin DH. 1997. The clinical diagnosis of genital ulcer disease in men. *Clin Infect Dis* 25:292–298. <http://dx.doi.org/10.1086/514548>.
- Leland DS, Ginocchio CC. 2007. Role of cell culture for virus detection in the age of technology. *Clin Microbiol Rev* 20:49–78. <http://dx.doi.org/10.1128/CMR.00002-06>.
- Proffitt MR, Schindler SA. 1995. Rapid detection of HSV with an enzyme-linked virus inducible system (ELVIS) employing a genetically modified cell line. *Clin Diagn Virol* 4:175–182. [http://dx.doi.org/10.1016/0928-0197\(95\)00011-V](http://dx.doi.org/10.1016/0928-0197(95)00011-V).
- Stabell EC, O'Rourke SR, Storch GA, Olivo PD. 1993. Evaluation of a genetically engineered cell line and a histochemical beta-galactosidase assay to detect herpes simplex virus in clinical specimens. *J Clin Microbiol* 31:2796–2798.
- Crist GA, Langer JM, Woods GL, Procter M, Hillyard DR. 2004. Evaluation of the ELVIS plate method for the detection and typing of herpes simplex virus in clinical specimens. *Diagn Microbiol Infect Dis* 49:173–177. <http://dx.doi.org/10.1016/j.diagmicrobio.2004.03.008>.
- LaRocco MT. 2000. Evaluation of an enzyme-linked viral inducible system for the rapid detection of herpes simplex virus. *Eur J Clin Microbiol Infect Dis* 19:233–235. <http://dx.doi.org/10.1007/s100960050466>.

16. Anderson NW, Buchan BW, Ledebner NA. 2014. Light microscopy, culture, molecular, and serologic methods for detection of herpes simplex virus. *J Clin Microbiol* 52:2–8. <http://dx.doi.org/10.1128/JCM.01966-13>.
17. Kimberlin DW, Rouse DJ. 2004. Clinical practice. Genital herpes. *N Engl J Med* 350:1970–1977. <http://dx.doi.org/10.1056/NEJMc023065>.
18. Wald A, Huang ML, Carrell D, Selke S, Corey L. 2003. Polymerase chain reaction for detection of herpes simplex virus (HSV) DNA on mucosal surfaces: comparison with HSV isolation in cell culture. *J Infect Dis* 188:1345–1351. <http://dx.doi.org/10.1086/379043>.
19. Baron JM, Rubben A, Grussendorf-Conen EI. 1996. Evaluation of a new general primer pair for rapid detection and differentiation of HSV-1, HSV-2, and VZV by polymerase chain reaction. *J Med Virol* 49:279–282. [http://dx.doi.org/10.1002/\(SICI\)1096-9071\(199608\)49:4<279::AID-JMV4>3.0.CO;2-0](http://dx.doi.org/10.1002/(SICI)1096-9071(199608)49:4<279::AID-JMV4>3.0.CO;2-0).
20. Shoji H, Koga M, Kusahara T, Kaji M, Ayabe M, Hino H, Hondo R. 1994. Differentiation of herpes simplex virus 1 and 2 in cerebrospinal fluid of patients with HSV encephalitis and meningitis by stringent hybridization of PCR-amplified DNAs. *J Neurol* 241:526–530. <http://dx.doi.org/10.1007/BF00873514>.
21. Selvaraju SB, Wurst M, Horvat RT, Selvarangan R. 2009. Evaluation of three analyte-specific reagents for detection and typing of herpes simplex virus in cerebrospinal fluid. *Diagn Microbiol Infect Dis* 63:286–291. <http://dx.doi.org/10.1016/j.diagmicrobio.2008.11.013>.
22. Koenig M, Reynolds KS, Aldous W, Hickman M. 2001. Comparison of Light-Cycler PCR, enzyme immunoassay, and tissue culture for detection of herpes simplex virus. *Diagn Microbiol Infect Dis* 40:107–110. [http://dx.doi.org/10.1016/S0732-8893\(01\)00260-7](http://dx.doi.org/10.1016/S0732-8893(01)00260-7).
23. Granato PA, Alkins BR, Yen-Lieberman B, Greene WH, Connolly J, Buchan BW, Ledebner NA. 2015. Comparative evaluation of AmpliVue HSV 1+2 assay with ELVIS culture for detecting herpes simplex virus 1 (HSV-1) and HSV-2 in clinical specimens. *J Clin Microbiol* 53:3922–3925.
24. Miller NS, Yen-Lieberman B, Poulter MD, Tang YW, Granato PA. 2012. Comparative clinical evaluation of the IsoAmp HSV assay with ELVIS HSV culture/ID/typing test system for the detection of herpes simplex virus in genital and oral lesions. *J Clin Virol* 54:355–358. <http://dx.doi.org/10.1016/j.jcv.2012.04.004>.
25. Ratnam S, Severini A, Zahariadis G, Petric M, Romanowski B. 2007. The diagnosis of genital herpes-beyond culture: an evidence-based guide for the utilization of polymerase chain reaction and herpes simplex virus type-specific serology. *Can J Infect Dis Med Microbiol* 18:233–240. <http://dx.doi.org/10.1155/2007/505364>.
26. Lafferty WE, Krofft S, Remington M, Giddings R, Winter C, Cent A, Corey L. 1987. Diagnosis of herpes simplex virus by direct immunofluorescence and viral isolation from samples of external genital lesions in a high-prevalence population. *J Clin Microbiol* 25:323–326.
27. Moseley RC, Corey L, Benjamin D, Winter C, Remington ML. 1981. Comparison of viral isolation, direct immunofluorescence, and indirect immunoperoxidase techniques for detection of genital herpes simplex virus infection. *J Clin Microbiol* 13:913–918.