



Identification of an Unusual 16S rRNA Mutation in *Neisseria gonorrhoeae*

Christine J. D. Guglielmino,^a Sally Appleton,^b Renu Vohra,^b Amy V. Jennison^a

^aPublic Health Microbiology, Queensland Health Forensic and Scientific Services, Brisbane, Australia

^bMicrobiology, QML Pathology, Brisbane, Australia

KEYWORDS *Neisseria gonorrhoeae*, gonorrhea, PCR, mutation, RNA, ribosomal, 16S, pharynx, false-negative reactions, Australia

Australian guidelines recommend screening and supplemental assays directed at different targets for molecular detection of *Neisseria gonorrhoeae*, and culture as the preferred test for nongenital sites despite its low sensitivity (1). QML Pathology uses the Aptima Combo 2 assay for screening, with Aptima GC for supplemental testing (Hologic, Marlborough, MA), both of which target the same 16S rRNA gene; the combined use of which the manufacturer considers appropriate because the assays target different regions (2). Cross-reactivity with other *Neisseria* spp. and false-negative reactions have been reported for non-16S rRNA gene-based *N. gonorrhoeae* gene targets (1). In October 2018, a 31-year-old male patient had a routine sexually transmitted infection (STI) screen performed on a throat swab, rectal swab, and urine for *Chlamydia trachomatis* and *N. gonorrhoeae*. The Aptima Combo 2 assay was positive for *N. gonorrhoeae* (RLU 1244), but the Aptima *N. gonorrhoeae* (GC) assay was equivocal (RLU 74) on the throat swab. Throat swabs were collected one week later and submitted for both nucleic acid amplification testing (NAAT) and culture. NAAT by the Aptima Combo 2 assay on the second throat swab was positive (RLU 1331); however, the Aptima *N. gonorrhoeae* (GC) assay was negative (RLU 26). Culture of the second throat swab grew *N. gonorrhoeae* (strain 9987), which was confirmed by phenotypic tests and Vitek MS (bioMérieux, France). The culture isolate was positive for *N. gonorrhoeae* with the Aptima Combo 2 assay (RLU 1290) and by Aptima GC assay (RLU 469). Both the urine and the rectal swab were negative for *N. gonorrhoeae* in the Aptima Combo 2 assay. The strain was whole-genome sequenced on the Illumina NextSeq 500 platform (San Diego, CA), and reads were assembled with Spades (3) (pubMLST identifier [ID] 84143). The strain was confirmed as *N. gonorrhoeae* by Kraken (4) and *rplF* sequence analysis (5) and typed as multilocus sequence type 1893 (MLST-1893), NgSTAR 142, NgMAST 8517; all preexisting profiles in their respective databases. The 16S rRNA sequence showed a single nucleotide polymorphism (G478T) to known sequences for this organism, which was not present in any sequences in the NCBI nucleotide database. This mutation lies within a sequence region listed in the US7172863B1 patent (6), causing an oligonucleotide binding mismatch likely responsible for the negative and equivocal reactions in the supplemental Aptima GC assay. The prevalence of this mutation in the local population is unknown, where 28% of gonococcal notifications have an isolate grown in culture (7), and sequencing of isolates is not routinely performed. Infections may potentially be underreported in areas reliant on this molecular assay for testing. There are three other strains currently in pubMLST with this same 16S sequence (56770, 56772, and 56774), all MLST-1893 and isolated in Spain in 2016, suggesting international dissemination of this strain. Other 16S mutations are associated with spectinomycin resistance (8); however, this strain did not demonstrate spectinomycin resistance (MIC < 64 µg/ml). This report highlights some of the ongoing challenges around the

Citation Guglielmino CJD, Appleton S, Vohra R, Jennison AV. 2019. Identification of an unusual 16S rRNA mutation in *Neisseria gonorrhoeae*. J Clin Microbiol 57:e01337-19. <https://doi.org/10.1128/JCM.01337-19>.

Editor Erik Munson, Marquette University

© Crown copyright 2019. The government of Australia, Canada, or the UK ("the Crown") owns the copyright interests of authors who are government employees. The [Crown Copyright](#) is not transferable.

Address correspondence to Christine J. D. Guglielmino, christine.guglielmino@health.qld.gov.au.

Accepted manuscript posted online 11 September 2019

Published 23 October 2019

selection of molecular diagnostic targets and multitest algorithms for gonorrhea and an advantage of using culture for nongenital site diagnosis.

Accession number(s). The whole-genome sequence has been deposited under SRA accession [PRJNA507689](#). The single nucleotide polymorphism has been deposited in GenBank under reference number [MK226480](#).

REFERENCES

1. Tapsall J, Lum G, Smith D. 2005. Guidelines for the use and interpretation of nucleic acid detection tests for *Neisseria gonorrhoeae* in Australia: a position paper on behalf of the Public Health Laboratory Network. *Commun Dis Intell Q Rep* 29:358–365.
2. Boyadzhyan B, Yashina T, Yatabe J, Patnaik M, Hill C. 2004. Comparison of the APTIMA CT and GC assays with the APTIMA combo 2 assay, the Abbott LCx assay, and direct fluorescent-antibody and culture assays for detection of *Chlamydia trachomatis* and *Neisseria gonorrhoeae*. *J Clin Microbiol* 42: 3089–3093. <https://doi.org/10.1128/JCM.42.7.3089-3093.2004>.
3. Bankevich A, Nurk S, Antipov D, Gurevich AA, Dvorkin M, Kulikov AS, Lesin VM, Nikolenko SI, Pham S, Pribelski AD, Pyshkin AV, Sirotkin AV, Vyahhi N, Tesler G, Alekseyev MA, Pevzner PA. 2012. SPAdes: a new genome assembly algorithm and its applications to single-cell sequencing. *J Comput Biol* 19:455–477. <https://doi.org/10.1089/cmb.2012.0021>.
4. Wood DE, Salzberg SL. 2014. Kraken: ultrafast metagenomic sequence classification using exact alignments. *Genome Biol* 15:R46. <https://doi.org/10.1186/gb-2014-15-3-r46>.
5. Bennett JS, Watkins ER, Jolley KA, Harrison OB, Maiden M. 2014. Identifying *Neisseria* species using the 50S ribosomal protein L6 (*rplF*) gene. *J Clin Microbiol* 52:1375–1381. <https://doi.org/10.1128/JCM.03529-13>.
6. Hogan JJ, Kop JA, McDonough SH. February 2007. Nucleic acid probes and methods for detecting *Neisseria gonorrhoeae*. U.S. patent US7172863B1.
7. Lahra MM, Enriquez R, George CR. 2019. Australian Gonococcal Surveillance Programme annual report, 2017. *Commun Dis Intell* (2018) 43:13. <https://doi.org/10.33321/cdi.2019.43.13>.
8. Galimand M, Gerbaud G, Courvalin P. 2000. Spectinomycin resistance in *Neisseria* spp. due to mutations in 16S rRNA. *Antimicrob Agents Chemother* 44:1365. <https://doi.org/10.1128/aac.44.5.1365-1366.2000>.