



Test Accuracy of Human Papillomavirus in Urine for Detection of Cervical Intraepithelial Neoplasia

Eliane Rohner,^a Lisa Rahangdale,^{b,c} Busola Sanusi,^d Andrea K. Knittel,^b Laurence Vaughan,^e Kirsty Chesko,^e Brian Faherty,^e Samantha E. Tulenko,^a John W. Schmitt,^f LaHoma S. Romocki,^g Vijay Sivaraman,^h Julie A. E. Nelson,ⁱ Jennifer S. Smith^{a,c}

^aDepartment of Epidemiology, Gillings School of Global Public Health, The University of North Carolina at Chapel Hill, Chapel Hill, North Carolina, USA

^bDepartment of Obstetrics and Gynecology, School of Medicine, The University of North Carolina at Chapel Hill, Chapel Hill, North Carolina, USA

^cLineberger Comprehensive Cancer Center, The University of North Carolina at Chapel Hill, Chapel Hill, North Carolina, USA

^dDepartment of Biostatistics, Gillings School of Global Public Health, The University of North Carolina at Chapel Hill, Chapel Hill, North Carolina, USA

^eBecton, Dickinson and Company, Sparks, Maryland, USA

^fDepartment of Obstetrics and Gynecology, Duke University School of Medicine, Durham, North Carolina, USA

^gDepartment of Public Health Education, North Carolina Central University, Durham, North Carolina, USA

^hDepartment of Biological and Biomedical Sciences, North Carolina Central University, Durham, North Carolina, USA

ⁱDepartment of Microbiology and Immunology, School of Medicine, The University of North Carolina at Chapel Hill, Chapel Hill, North Carolina, USA

Eliane Rohner and Lisa Rahangdale contributed equally to this work. Author order was determined in order of increasing seniority.

ABSTRACT The objective was to assess the diagnostic test accuracy of high-risk human papillomavirus (hrHPV) testing of self-collected urine and cervicovaginal samples for the detection of cervical intraepithelial neoplasia grade 2 or higher (CIN2+). We recruited a convenience sample of women 25 to 65 years of age who were undergoing clinically indicated colposcopy at two medical centers in North Carolina between November 2016 and January 2019. Women with normal cytology results and positive hrHPV results were also recruited. Urine samples, self-collected cervicovaginal samples, provider-collected cervical samples, and cervical biopsy samples were obtained from all enrolled women. Samples were tested for hrHPV DNA using the Onclarity assay (Becton Dickinson, Sparks, MD). Biopsy samples were histologically graded as CIN2+ or <CIN2. We calculated the sensitivity and specificity for detection of CIN2+ and assessed agreement between sample collection methods. We included 307 women (median age, 36 years) with valid histology results and triple-matched urine, self-collected cervicovaginal, and provider-collected cervical hrHPV results; 83 women (27%) had CIN2+. Urine-based hrHPV testing correctly identified 80% of CIN2+ cases (95% confidence interval [CI], 71 to 88%) using the PCR cycle threshold (C_T) established for provider-collected cervical samples, but sensitivity remained below the estimates for self-collected cervicovaginal and provider-collected cervical samples (both 94% [95% CI, 89 to 99%]). Using a higher C_T cutoff value of ≤ 40 , 90% sensitivity was achieved for urine-based hrHPV testing. Agreement between results for urine samples and self-collected cervicovaginal samples ($\kappa = 0.58$) or provider-collected cervical samples ($\kappa = 0.54$) was moderate. Urine-based hrHPV testing may be a promising approach to improve cervical cancer screening coverage, especially among women with limited access to health care.

KEYWORDS HPV testing, cervical intraepithelial neoplasia, diagnostic test accuracy, human papillomavirus, self-collection, urine

Although invasive cervical cancer (ICC) is preventable through human papillomavirus (HPV) vaccination, as well as screening and treatment of precancerous cervical lesions, the global ICC burden remains high (1). U.S. guidelines currently recommend screening with cytology alone, primary high-risk HPV (hrHPV) testing, or cytology-

Citation Rohner E, Rahangdale L, Sanusi B, Knittel AK, Vaughan L, Chesko K, Faherty B, Tulenko SE, Schmitt JW, Romocki LS, Sivaraman V, Nelson JAE, Smith JS. 2020. Test accuracy of human papillomavirus in urine for detection of cervical intraepithelial neoplasia. *J Clin Microbiol* 58:e01443-19. <https://doi.org/10.1128/JCM.01443-19>.

Editor Yi-Wei Tang, Cepheid

Copyright © 2020 American Society for Microbiology. All Rights Reserved.

Address correspondence to Jennifer S. Smith, jennifers@unc.edu.

Received 1 September 2019

Returned for modification 30 September 2019

Accepted 12 December 2019

Accepted manuscript posted online 2 January 2020

Published 24 February 2020

hrHPV cotesting every 3 to 5 years, depending on the woman's age (2–5). Unfortunately, not all women are screened at the recommended intervals, and most ICC cases occur among underscreened women (6–8). Barriers to screening include cost, inflexible working hours, lack of transportation and childcare, fear of pain, and embarrassment (9, 10). New screening approaches addressing these barriers are needed to improve cervical cancer prevention.

While cytology requires a provider-collected cervical sample, hrHPV testing can be performed on self-collected specimens. Noninvasive, urine-based hrHPV testing might be especially attractive for women who are reluctant to undergo gynecological examinations or to perform cervicovaginal self-sampling. For the detection of cervical intraepithelial neoplasia grade 2 or higher (CIN2+), self-collected cervicovaginal samples and provider-collected samples for hrHPV testing are similarly sensitive (11, 12). The performance of urine-based hrHPV testing for CIN2+ detection is less clear, because studies have used various hrHPV assays and nonstandardized urine collection methods (13–22).

Our objective was to assess the accuracy of urine-based hrHPV testing for CIN2+ detection using media to preserve urine samples. We compared the accuracy of urine-based hrHPV testing to testing of self-collected cervicovaginal and provider-collected cervical samples and assessed the agreement between different sampling methods.

(Preliminary results of this work were presented at EUROGIN 2018, Lisbon, Portugal, 2 to 5 December 2018.)

MATERIALS AND METHODS

Study population. Between November 2016 and January 2019, we recruited a convenience sample of women, 25 to 65 years of age, who were attending colposcopy clinics at the University of North Carolina at Chapel Hill (UNC) Women's Hospital or Duke University Hospital for one of the following indications: abnormal cytology results, infection with HPV-16/18, persistent infection with other hrHPV subtypes, or treatment for CIN2+. We also aimed to include a sample of HPV-positive women at lower risk for CIN2+, who would be more similar to a primary screening population than the patients undergoing clinically indicated colposcopy. Therefore, we invited women with normal cytology findings and positive hrHPV results for HPV types other than HPV-16/18 at the time of their routine screening. This group was referred to as "research only," because current U.S. guidelines do not recommend immediate referral for colposcopy for such women (4).

Potentially eligible women were identified through review of electronic medical records and were contacted via telephone or during their clinic visits. Women were excluded if they were pregnant or had had their cervix removed; additionally, women in the research only group were excluded if they were taking anticoagulants or if the enrollment date was not within 3 months after their original hrHPV diagnosis. Written informed consent was obtained from all participating women. The study was approved by the institutional review boards at UNC and Duke University.

Sample collection. During the clinic visit, participating women received detailed verbal and written instructions, in English or Spanish, concerning the study procedures. The women then provided two urine samples, i.e., an initial-stream sample of ~20 ml for hrHPV testing and a midstream sample of up to 100 ml for pregnancy testing, if clinically indicated. Next, the women self-collected a cervicovaginal sample by inserting a Viba brush (Rovers Medical Devices BV, Oss, The Netherlands) to the top of the vaginal canal, rotating it five times, and releasing the brush head into a vial prefilled with 6 ml of preservative liquid-based cytology medium (ThinPrep; Hologic Inc., Bedford, MA).

Next, a pelvic examination was performed by a clinical provider, during which a cervical scraping was collected with two 360° turns, in a clockwise fashion, of a brush-like cervical cell collector (Wallach Papette; Wallach Surgical Devices, Trumbull, CT). The provider-collected cervical sample was preserved in a 20-ml vial of ThinPrep medium for subsequent hrHPV testing. Colposcopy was performed for all women, following cervical treatment with 3 to 5% acetic acid (followed by Lugol's iodine at the Duke University Hospital site), according to standard clinical procedures. Directed biopsy samples were taken from visible cervical lesions, and endocervical curettage (ECC) was performed if the transformation zone or the limits of a lesion near the cervical os could not be fully visualized. If no cervical lesions were observable, then one random biopsy sample was taken at the 12 o'clock position of the cervix and ECC was performed. At least two cervical biopsy samples were obtained during each procedure. The loop electrosurgical excision procedure was performed when clinically indicated. At the end of the visit, the women received a gift card (\$40 for women who had undergone clinically indicated colposcopy and \$120 for research only participants).

Sample processing and laboratory analyses. From the initial-stream specimen, 2 ml of urine was transferred (within 5 min after collection) into a Becton Dickinson (BD) molecular tube containing 0.2 ml of a proprietary preservative to ensure sample stability. The tubes are fitted with a pierceable cap to facilitate automated sample processing on the BD Viper LT System. All samples and the BD molecular

tube were placed in a cooler with frozen gel packs within 10 min after sample collection and were kept cool until they could be further aliquoted the same day. The self- and provider-collected samples were vortex-mixed for 10 to 30 s, and 0.5 ml of each sample was transferred to separate BD molecular tubes containing 1.7 ml of an HPV diluent buffer. All three BD molecular tubes were stored at -20°C until they were shipped to BD for hrHPV testing using the Onclarity assay (BD, Sparks, MD). The staff at BD did not have access to any clinical information concerning the participants. The Onclarity assay uses PCR and nucleic acid hybridization to detect DNA of 14 hrHPV subtypes. Six hrHPV subtypes (subtypes 16, 18, 31, 45, 51, and 52) are individually genotyped, and the other 8 subtypes are identified in three groups (subtypes 33/58, 56/59/66, and 35/39/68). Provider- and self-collected Hologic ThinPrep specimens were processed using the standard liquid-based cytology workflow on the BD Viper LT System. The Onclarity assay has been approved by the FDA for use with provider-collected samples in BD SurePath preservative fluid, but it shows similar performance when used with ThinPrep PreservCyt transport medium (23). Urine samples were processed using a co-collection device workflow, without a prewarming step. The remaining provider-collected cervical sample was sent to the UNC cytopathology laboratory for cytological analysis, if clinically indicated, or storage. The Onclarity hrHPV test results were obtained for research purposes only and were not shared with participants.

Histological diagnoses ($<\text{CIN}2$ versus $\text{CIN}2+$) served as the reference standard for this test accuracy study. Cervical biopsy samples from women who underwent colposcopy as part of their scheduled clinical appointments were sent to the UNC or Duke University Hospital surgical pathology laboratory for histological evaluation, according to standard procedures. Pathologists had access to clinical information captured in the electronic medical records but were unaware of the study hrHPV results. The colposcopy patients were informed of the histological results by the clinical team responsible for their care. Biopsy samples taken from women in the research only group were analyzed by a gynecological pathologist at the UNC translational pathology laboratory, who did not have access to clinical information. Women in the research only group were contacted with histological results by the study team after review by a practicing gynecologist. Women whose samples were inadequate for pathology review were invited to return for additional sampling. Women with $\text{CIN}2+$ were referred for further treatment, according to the standard of care.

Sample size considerations, definitions, and statistical analyses. Study recruitment was guided by an expected $\text{CIN}2+$ prevalence of 31% among the colposcopy clinic patients (13) and 6% among the research only participants (24). Our sample size calculations were focused on estimating the sensitivity of urine-based hrHPV testing with adequate precision. We expected 90% sensitivity and aimed to achieve sampling error margins of 10 percentage points. Thus, we aimed to recruit 160 colposcopy patients (49 $\text{CIN}2+$ cases expected) and 250 research only participants (15 $\text{CIN}2+$ cases expected). However, recruitment of research only participants turned out to be logistically challenging, with 79% of eligible women declining participation, cancelling clinic visits, or not being reachable. Therefore, we recruited more colposcopy patients and fewer research only participants than planned.

We included women with valid triple-matched results (urine, self-collected cervicovaginal, and provider-collected cervical samples available) in the test accuracy analysis; participants with missing samples or invalid test results were excluded. Participants without a valid cervical biopsy sample result (e.g., due to insufficient tissue) were also excluded. There is no established PCR cycle threshold (C_T) for the Onclarity assay in urine; therefore, we used the same cutoff value as for provider-collected cervical samples (≤ 34.2 for all hrHPV subtypes except HPV-16, with a C_T of ≤ 38.3). In an exploratory analysis, we evaluated the effects of different C_T cutoff values (≤ 36.0 , ≤ 38.3 , and ≤ 40.0) on the accuracy of urine-based hrHPV testing for $\text{CIN}2+$. Samples that were reactive for any hrHPV subtypes detected by the Onclarity assay were regarded as positive.

We used descriptive statistics to assess the sociodemographic characteristics of included women. We calculated the sensitivity and specificity of hrHPV testing for $\text{CIN}2+$ detection for different sample types. McNemar's test was used to assess whether the sensitivity of hrHPV testing with urine samples was different from that with self-collected cervicovaginal samples and provider-collected cervical samples. The level of statistical significance was set at <0.05 . We computed the overall agreement and the specific positive agreement and negative agreement between hrHPV test results for urine samples and those for self-collected cervicovaginal samples or provider-collected cervical samples (25). We calculated the unweighted Cohen's kappa and its 95% confidence interval (CI) to assess the degree of agreement beyond chance between the different sample types. The following interpretation was used: ≤ 0 , no agreement; 0.01 to 0.20, slight agreement; 0.21 to 0.40, fair agreement; 0.41 to 0.60, moderate agreement; 0.61 to 0.80, substantial agreement; 0.81 to 1.00, almost perfect agreement (26).

In sensitivity analyses, we examined whether the test accuracy and agreement results changed if we restricted the study population to patients with clinically indicated colposcopy or lower-risk research only participants. We also computed test accuracy with all valid results for a given sample type (403 results for urine samples, 380 results for cervicovaginal samples, and 343 results for provider-collected cervical samples), and we estimated hrHPV test accuracy for $\text{CIN}3+$ detection. Analyses were performed using SAS/STAT software (SAS Institute Inc., Cary, NC).

Data availability. Study data will be shared upon request, in accordance with the National Institutes of Health (NIH) guidelines on data sharing, as proposed in our original grant submission.

RESULTS

Study population. A total of 434 women (363 with clinically indicated colposcopy and 71 in the research only group) were enrolled (Fig. 1). Of those women, 413 women had valid histology results. In the test accuracy analysis, we included 307 women with




FIG 1 Flow diagram showing the number of women eligible for inclusion during the study period and the number of women excluded from the analysis, with reasons.

valid histology results and valid results for their triple-matched urine, self-collected cervicovaginal, and provider-collected cervical samples. The median age of these women was 36 years (interquartile range [IQR], 31 to 45 years). The study population was ethnically and racially diverse, with 38% non-Hispanic white women ($n = 117$), 29% Hispanic women ($n = 90$), and 26% non-Hispanic black women ($n = 79$) (Table 1). Participant health insurance coverage ranged from private ($n = 118$ [39%]) or government-sponsored ($n = 70$ [23%]) to no health insurance ($n = 116$ [38%]).

Diagnostic test accuracy for CIN2+ detection. Of 307 included women, 83 (27%) had histologically confirmed CIN2+ (34 CIN2 cases, 11 CIN2/3 cases, 36 CIN3 cases, and 2 ICC cases) and 224 (73%) had <CIN2 (57 CIN1 cases and 167 cases without dysplasia). The most common hrHPV subtypes detected in urine samples were HPV-16, with 58 positive samples (19%), HPV-56/59/66, with 56 samples positive for at least one of these subtypes (18%), and HPV-35/39/68, with 35 samples positive for at least one of these subtypes (11%). HPV-18 was detected in 17 urine samples (6%), and HPV-45 was detected in 18 urine samples (6%). Urine-based hrHPV testing correctly identified 66 of 83 CIN2+ cases when the C_7 cutoff value for provider-collected samples was used. The sensitivity of hrHPV testing in urine samples (80.0% [95% CI, 70.8 to 88.2%]) was significantly lower than that in self-collected cervicovaginal samples or provider-collected cervical samples (both 94.0% [95% CI, 88.9 to 99.1%]; $P = 0.001$). The specificity of hrHPV testing was the same in urine samples and provider-collected cervical

TABLE 1 Baseline characteristics of women included in the test accuracy analysis, with valid urine samples, self-collected cervicovaginal samples, and provider-collected cervical samples

Parameter	Result(s) (n = 307)
Age (median [IQR]) (yr)	36 (31–45)
Race/ethnicity (no. [%])	
Hispanic	90 (29)
Non-Hispanic white	117 (38)
Non-Hispanic black	79 (26)
Other ^a	20 (7)
Missing data	1
Marital status (no. [%])	
Married/living with partner	117 (40)
Divorced/separated	69 (23)
Widowed	10 (3)
Single	100 (34)
Missing data	11
Education (no. [%])	
Elementary school or less	27 (9)
High school	90 (30)
Some college	103 (35)
College graduate	77 (26)
Missing data	10
Monthly income (median [IQR]) (\$)	2,200 (1,456–4,000)
Unemployed	10
Missing data	44
Health insurance (no. [%])	
Private	118 (39)
Medicaid/Medicare/TRICARE	70 (23)
None	116 (38)
Missing data	3
No. of live births (median [IQR])	2 (1–3)
Missing data	6
No. of sex partners in past 3 mo (median [IQR])	1 (1–1)
Missing data	3
Current smoker (no. [%])	68 (22)
Missing data	4

^aOther races included American Indian/Alaskan Native (n = 4), Asian (n = 9), black-Indian (n = 1), white-black (n = 1), white-Asian (n = 2), Indian (n = 1), Mediterranean (n = 1), and Native Hawaiian/other Pacific Islander (n = 1).

samples (38.4% [95% CI, 32.0 to 44.8%]), but it was lower in self-collected cervicovaginal samples (30.0% [95% CI, 23.9 to 35.9%]; *P* = 0.004) (Table 2).

Table 3 shows test accuracy results for urine-based hrHPV testing using different *C*₇ cutoff values. When we chose a higher *C*₇ cutoff value of ≤40.0, we correctly identified an additional 9 CIN2+ cases and the sensitivity estimate increased to 90.4% (95% CI, 84.0 to 96.7%), which was similar to the sensitivity estimates for self-collected cervicovaginal samples (*P* = 0.37) and provider-collected cervical samples (*P* = 0.45). However, at this cutoff value the assay also yielded positive hrHPV results for an additional 37 <CIN2 cases and the specificity thus decreased to 21.9% (95% CI, 16.5 to 27.3%).

Agreement between hrHPV results in different samples. Agreement results are based on the *C*₇ cutoff value for provider-collected samples. Overall agreement between hrHPV results in urine and self-collected cervicovaginal samples was 83% (95% CI, 79 to 87%), with positive agreement of 88% (95% CI, 85 to 91%) and negative agreement of 70% (95% CI, 62 to 77%); agreement beyond chance was moderate (kappa = 0.58 [95% CI, 0.48 to 0.68]). Results were similar for the agreement between hrHPV results in urine and provider-collected cervical samples (Table 4).

TABLE 2 Diagnostic test accuracy of the Onclarity hrHPV assay for detection of CIN2+ with different sample types, using the Onclarity C_7 cutoff value established for provider-collected samples for all sample types

Sample type and hrHPV assay result ^a	No. with histology result of:		Sensitivity (95% CI) (%)	Specificity (95% CI) (%)
	CIN2+ (n = 83)	<CIN2 (n = 224)		
Urine samples				
Positive	66	138	80.0 (70.8–88.2)	38.4 (32.0–44.8)
Negative	17	86		
Self-collected cervicovaginal samples				
Positive	78	157	94.0 (88.9–99.1)	30.0 (23.9–35.9)
Negative	5	67		
Provider-collected cervical samples				
Positive	78	138	94.0 (88.9–99.1)	38.4 (32.0–44.8)
Negative	5	86		

^aA C_7 cutoff value of ≤ 34.2 was used for all hrHPV subtypes except for HPV-16, for which the C_7 cutoff value was set at ≤ 38.3 , according to FDA-approved thresholds for provider-collected specimens.

Sensitivity analyses. When we restricted the analyses to 243 patients with clinically indicated colposcopy and a CIN2+ prevalence of 33% (81 CIN2+ cases), the results were similar (see Table S1 in the supplemental material). Agreement between hrHPV test results in urine samples and self-collected cervicovaginal samples or provider-collected cervical samples remained moderate (Table S2). Results limited to the 64 research only participants were imprecise, due to the small sample size and the small number of CIN2+ cases ($n = 2$ [3%]) (Tables S3 and S4).

When we included all valid results for a given sample type in the test accuracy analysis (403 results for urine samples, 380 results for self-collected cervicovaginal samples, and 343 results for provider-collected cervical samples) (Table S5), the results were similar to those of the main analysis. When we defined CIN3+ as the target condition, the sensitivity estimates increased marginally (Table S6). Urine-based hrHPV testing was 83.7% sensitive (95% CI, 73.3 to 94.0%) for CIN3+ detection.

DISCUSSION

We found that urine-based Onclarity hrHPV testing correctly identified 80% of CIN2+ cases using the C_7 cutoff value for provider-collected cervical samples, but sensitivity remained below the estimates for self-collected cervicovaginal samples and provider-collected samples (both 94%). When we changed the C_7 cutoff value for urine samples to ≤ 40 , 90% sensitivity was achieved. Using the C_7 cutoff value for provider-collected samples, agreement between hrHPV results for urine samples and those for self-collected cervicovaginal samples or provider-collected cervical samples was moderate.

To our knowledge, this is the first study assessing the accuracy of urine-based hrHPV testing for CIN2+ detection using the BD Onclarity assay. Several studies have examined the accuracy of urine-based hrHPV testing for CIN2+ detection using other assays (13, 14, 16–22, 27–29), and most studies have focused on the analytical accuracy of urine-based testing for detection of hrHPV infection (30). Test accuracy estimates vary substantially between studies. An early study evaluated the Hybrid Capture II assay and found low sensitivity of 45% and specificity of 70% for urine hrHPV testing for CIN2+ detection (16). Our findings are in line with later studies that showed higher sensitivities

TABLE 3 Test accuracy of the urine-based Onclarity hrHPV assay for detection of CIN2+ using different assay cutoff values

C_7 cutoff value ^a	Sensitivity (95% CI) (%)	Specificity (95% CI) (%)
≤ 40.0	90.4 (84.0–96.7)	21.9 (16.5–27.3)
≤ 38.3	88.0 (81.0–95.0)	26.3 (20.6–32.1)
≤ 36.0	80.7 (72.2–89.2)	32.6 (26.5–38.7)

^aThe same C_7 cutoff value was used for all hrHPV subtypes detected by the Onclarity assay.

TABLE 4 Agreement between Onclarity hrHPV test results for urine samples and those for self-collected cervicovaginal samples or provider-collected cervical samples

Sample type and hrHPV assay result	Urine sample result (no. [%])		Agreement (95% CI) (%)			Cohen's kappa (95% CI)
	Positive	Negative	Overall	Positive	Negative	
	Self-collected cervicovaginal samples					
Positive	193 (63)	42 (14)	83 (79–87)	88 (85–91)	70 (62–77)	0.58 (0.48–0.68)
Negative	11 (4)	61 (20)				
Provider-collected cervical samples						
Positive	179 (58)	37 (12)	80 (75–84)	85 (82–89)	68 (61–76)	0.54 (0.43–0.64)
Negative	25 (8)	66 (22)				

of 80 to 100% and low specificities of 25 to 53% among patients referred for colposcopy, a population at higher risk of HPV infection and CIN2+ (13, 14, 18–21, 28, 29). The more recent studies, including ours, added preservation solutions to stabilize urine samples. One study used the Aptima HPV RNA test (Hologic, Inc.) and found a lower sensitivity of 45% and a specificity of 62% for CIN2+ (17). To date, most studies have been conducted among women referred for colposcopy, and evidence is very limited regarding the performance of urine-based hrHPV testing for CIN2+ detection in primary screening populations (27). The specificity of hrHPV testing is expected to be lower among colposcopy patients than in screening populations, as specificity decreases with increasing hrHPV prevalence (31).

In our study, the sensitivity of urine-based hrHPV testing for CIN2+ detection was relatively lower than that for self-collected cervicovaginal samples and provider-collected samples when the C_T cutoff value for provider-collected cervical samples was used. Currently, there is no established C_T cutoff value for urine-based Onclarity hrHPV testing. At a C_T cutoff value of ≤ 40 , the sensitivity of urine-based hrHPV testing increased to 90% for detection of CIN2+, but the specificity decreased to 22%. Additional investigations are needed to determine the optimal C_T cutoff value for the Onclarity assay for urine samples and to further increase the test accuracy of urine-based hrHPV testing. Of note, we attempted to optimize the urine collection process by taking initial-stream urine and quickly adding a preservative solution to stabilize the samples (32). However, the cup-based collection of initial-stream urine, followed by the addition of preservative, may not be as efficient as collecting initial-stream urine with a cup/device that immediately mixes it with preservative.

Strengths of our study include the relatively large sample size and standardized sample collection procedures. We instructed women to collect ~20 ml of initial-stream urine and then transferred a fixed amount of 2 ml into a BD molecular tube containing 0.2 ml of a proprietary preservative, to obtain similar dilutions for all samples. To avoid verification bias, we aimed to perform the reference test (histology) for all participants and included only women with valid histology results in the test accuracy analysis. We recruited both high-risk women referred for colposcopy and women with normal cytology results and hrHPV infection with other than subtypes 16/18 (the research only group). In a sensitivity analysis, we assessed whether results changed if we restricted our population to women referred for colposcopy; this was not the case. Unfortunately, recruitment of research only participants was logistically challenging, and we could not reach a large enough sample size to draw meaningful conclusions regarding urine validation for this group separately. Another limitation of our study is that the results we found in a CIN2+-enriched population are not necessarily generalizable to a primary screening population.

Although further work is needed to improve its accuracy, urine hrHPV testing has the potential to increase screening options for underscreened women (33). Cost-effectiveness analyses from a program perspective are also warranted, to determine whether urine-based hrHPV screening may be preferred over hrHPV testing based on provider-collected cervical samples or self-collected cervicovaginal samples. Noninva-

sive urine hrHPV samples seem to be preferred by women over both provider-collected cervical samples (13, 16) and self-collected cervicovaginal samples (13, 16, 22). Furthermore, urine samples could be collected at home and sent by mail to the laboratory for hrHPV testing, thereby removing the need for an initial clinic-based pelvic examination. In a recent study, hrHPV prevalence rates were similar for urine samples collected at home or in the clinic, and most participants were comfortable receiving a urine collection kit in the mail (13).

Urine-based hrHPV tests have not yet been approved by the FDA for clinical use, and further studies to evaluate and to improve the accuracy of urine-based hrHPV testing for CIN2+ detection in a primary screening population and among women living with HIV are needed. Nevertheless, urine-based hrHPV testing may be a viable option to improve cervical cancer screening coverage in the future, especially among women with limited access to health care.

SUPPLEMENTAL MATERIAL

Supplemental material is available online only.

SUPPLEMENTAL FILE 1, PDF file, 0.2 MB.

ACKNOWLEDGMENTS

We are grateful to all included women for their participation in the study. We thank Anna Baker, Johana Bravo, Elena DiRosa, Sara Smith, and Claire Edelman for the day-to-day coordination of this study, including the recruitment of participants and data collection, and the staff members of the UNC and Duke University colposcopy clinics who supported this study. We are grateful to Siobhan O'Connor for reviewing the histopathology slides for the research only participants. Furthermore, we thank the UNC Center for AIDS Research HIV/STD Laboratory Core for sample processing and storage and clinical trial support. We also thank BD for donating the Onclarity HPV testing kits, Hologic, Inc., for donating preservation media, cervical sample collection devices, and imager slides, and Rovers Medical Devices for donating self-collection brushes.

This research was supported by NIH grant U54 CA156733 to J.S.S. and the UNC Center for AIDS Research, an NIH-funded program (grant P30 AI50410). E.R. was supported by a grant from the Swiss Cancer Research Foundation (grant BIL KFS-4423-02-2018). Additional support to A.K.K. was provided by the Division of General Obstetrics and Gynecology Women's Health Research Fellowship. Specimen transport tubes and HPV testing were donated by BD, self-collection brushes were donated by Rovers Medical Devices BV, and cytology medium was donated by Hologic. The content of this publication is solely the responsibility of the authors and does not necessarily represent the official views of the funders.

J.S.S. has received research grants, supply donations, and consultancies from, served on paid advisory boards for, and/or been a paid speaker for Arbor Vita, BD, Hologic, Rovers Medical Devices, and Trovagene in the past 5 years. J.A.E.N. has received financial support from Hologic for work on a different study. L.V., K.C., and B.F. are employed by BD, which sells and distributes cervical cancer screening products, including the BD Onclarity HPV assay.

REFERENCES

1. Ferlay J, Colombet M, Soerjomataram I, Mathers C, Parkin DM, Piñeros M, Znaor A, Bray F. 2019. Estimating the global cancer incidence and mortality in 2018: GLOBOCAN sources and methods. *Int J Cancer* 144: 1941–1953. <https://doi.org/10.1002/ijc.31937>.
2. Curry SJ, US Preventive Services Task Force, Krist AH, Owens DK, Barry MJ, Caughey AB, Davidson KW, Doubeni CA, Epling JW, Kemper AR, Kubik M, Landefeld CS, Mangione CM, Phipps MG, Silverstein M, Simon MA, Tseng C-W, Wong JB. 2018. Screening for cervical cancer. *JAMA* 320:674–686. <https://doi.org/10.1001/jama.2018.10897>.
3. Committee on Practice Bulletins-Gynecology. 2016. Practice bulletin no. 168: cervical cancer screening and prevention. *Obstet Gynecol* 128: e111–e130. <https://doi.org/10.1097/AOG.0000000000001708>.
4. Saslow D, Solomon D, Lawson HW, Killackey M, Kulasingam SL, Cain J, Garcia FAR, Moriarty AT, Waxman AG, Wilbur DC, Wentzensen N, Downs LS, Spitzer M, Moscicki A-B, Franco EL, Stoler MH, Schiffman M, Castle PE, Myers ER. 2012. American Cancer Society, American Society for Colposcopy and Cervical Pathology, and American Society for Clinical Pathology screening guidelines for the prevention and early detection of cervical cancer. *CA Cancer J Clin* 62:147–172. <https://doi.org/10.3322/caac.21139>.
5. Huh WK, Ault KA, Chelmsow D, Davey DD, Goulart RA, Garcia FAR, Kinney WK, Massad LS, Mayeaux EJ, Saslow D, Schiffman M, Wentzensen N, Lawson HW, Einstein MH. 2015. Use of primary high-risk human papillomavirus testing for cervical cancer screening. *Obstet Gynecol* 125: 330–337. <https://doi.org/10.1097/AOG.0000000000000669>.

6. Andrae B, Kemetli L, Sparen P, Silfverdal L, Strander B, Ryd W, Dillner J, Tornberg S. 2008. Screening-preventable cervical cancer risks: evidence from a nationwide audit in Sweden. *J Natl Cancer Inst* 100:622–629. <https://doi.org/10.1093/jnci/djn099>.
7. Spence AR, Goggin P, Franco EL. 2007. Process of care failures in invasive cervical cancer: systematic review and meta-analysis. *Prev Med* 45:93–106. <https://doi.org/10.1016/j.ypmed.2007.06.007>.
8. Pruitt SL, Werner CL, Borton EK, Sanders JM, Balasubramanian BA, Barnes A, Betts AC, Skinner CS, Tiro JA. 2018. Cervical cancer burden and opportunities for prevention in a safety-net healthcare system. *Cancer Epidemiol Biomarkers Prev* 27:1398–1406. <https://doi.org/10.1158/1055-9965.EPI-17-0912>.
9. Ackerson K, Gretebeck K. 2007. Factors influencing cancer screening practices of underserved women. *J Am Acad Nurse Pract* 19:591–601. <https://doi.org/10.1111/j.1745-7599.2007.00268.x>.
10. Akinlotan M, Bolin JN, Helduser J, Ojinnaka C, Lichorad A, McClellan D. 2017. Cervical cancer screening barriers and risk factor knowledge among uninsured women. *J Community Health* 42:770–778. <https://doi.org/10.1007/s10900-017-0316-9>.
11. Arbyn M, Verdoodt F, Snijders PJF, Verhoef VMJ, Suonio E, Dillner L, Minozzi S, Bellisario C, Banzi R, Zhao F-H, Hillemanns P, Anttila A. 2014. Accuracy of human papillomavirus testing on self-collected versus clinician-collected samples: a meta-analysis. *Lancet Oncol* 15:172–183. [https://doi.org/10.1016/S1470-2045\(13\)70570-9](https://doi.org/10.1016/S1470-2045(13)70570-9).
12. Snijders PJF, Verhoef VMJ, Arbyn M, Ogilvie G, Minozzi S, Banzi R, Van Kemenade FJ, Heideman DAM, Meijer C. 2013. High-risk HPV testing on self-sampled versus clinician-collected specimens: a review on the clinical accuracy and impact on population attendance in cervical cancer screening. *Int J Cancer* 132:2223–2236. <https://doi.org/10.1002/ijc.27790>.
13. Senkomago V, Des Marais AC, Rahangdale L, Vibat CRT, Erlander MG, Smith JS. 2016. Comparison of urine specimen collection times and testing fractions for the detection of high-risk human papillomavirus and high-grade cervical precancer. *J Clin Virol* 74:26–31. <https://doi.org/10.1016/j.jcv.2015.11.005>.
14. Sahasrabudhe VV, Gravitt PE, Dunn ST, Robbins D, Brown D, Allen RA, Eby YJ, Smith KM, Zuna RE, Zhang RR, Gold MA, Schiffman M, Walker JL, Castle PE, Wentzensen N. 2014. Evaluation of clinical performance of a novel urine-based HPV detection assay among women attending a colposcopy clinic. *J Clin Virol* 60:414–417. <https://doi.org/10.1016/j.jcv.2014.04.016>.
15. Fambrini M, Penna C, Pieralli A, Bussani C, Fallani MG, Andersson KL, Scarselli G, Marchionni M. 2008. PCR detection rates of high risk human papillomavirus DNA in paired self-collected urine and cervical scrapes after laser CO₂ conization for high-grade cervical intraepithelial neoplasia. *Gynecol Oncol* 109:59–64. <https://doi.org/10.1016/j.ygyno.2007.12.032>.
16. Sellors JW, Lorincz AT, Mahony JB, Mielzynska I, Lytwyn A, Roth P, Howard M, Chong S, Daya D, Chapman W, Chernesky M. 2000. Comparison of self-collected vaginal, vulvar and urine samples with physician-collected cervical samples for human papillomavirus testing to detect high-grade squamous intraepithelial lesions. *CMAJ* 163:513–518.
17. Ascitto KC, Ernstson A, Forslund O, Borgfeldt C. 2018. Self-sampling with HPV mRNA analyses from vagina and urine compared with cervical samples. *J Clin Virol* 101:69–73. <https://doi.org/10.1016/j.jcv.2018.02.002>.
18. Bernal S, Palomares JC, Artura A, Parra M, Cabezas JL, Robles A, Martín Mazuelos E. 2014. Comparison of urine and cervical samples for detecting human papillomavirus (HPV) with the Cobas 4800 HPV test. *J Clin Virol* 61:548–552. <https://doi.org/10.1016/j.jcv.2014.10.001>.
19. Cuzick J, Cadman L, Ahmad AS, Ho L, Terry G, Kleeman M, Lyons D, Austin J, Stoler MH, Vibat CRT, Dockter J, Robbins D, Billings PR, Erlander MG. 2017. Performance and diagnostic accuracy of a urine-based human papillomavirus assay in a referral population. *Cancer Epidemiol Biomarkers Prev* 26:1053–1059. <https://doi.org/10.1158/1055-9965.EPI-16-0960>.
20. Leeman A, del Pino M, Molijn A, Rodriguez A, Torné A, de Koning M, Ordi J, van Kemenade F, Jenkins D, Quint W. 2017. HPV testing in first-void urine provides sensitivity for CIN2+ detection comparable with a smear taken by a clinician or a brush-based self-sample: cross-sectional data from a triage population. *BJOG* 124:1356–1363. <https://doi.org/10.1111/1471-0528.14682>.
21. Lorenzi AT, Fregnani JHT, Dockter J, Fitzgerald K, Strohecker E, Eaton B, Vibat CRT, Erlander MG, Scapulatempo-Neto C, Smith JS, Longatto-Filho A. 2018. High-risk human papillomavirus detection in urine samples from a referral population with cervical biopsy-proven high-grade lesions. *J Low Genit Tract Dis* 22:17–20. <https://doi.org/10.1097/LGT.0000000000000352>.
22. Sargent A, Fletcher S, Bray K, Kitchener HC, Crosbie EJ. 2019. Cross-sectional study of HPV testing in self-sampled urine and comparison with matched vaginal and cervical samples in women attending colposcopy for the management of abnormal cervical screening. *BMJ Open* 9:e025388. <https://doi.org/10.1136/bmjopen-2018-025388>.
23. Cuzick J, Ahmad AS, Austin J, Cadman L, Ho L, Terry G, Kleeman M, Ashdown-Barr L, Lyons D, Stoler M, Szarewski A. 2016. A comparison of different human papillomavirus tests in PreservCyt versus SurePath in a referral population-PREDICTORS 4. *J Clin Virol* 82:145–151. <https://doi.org/10.1016/j.jcv.2016.06.015>.
24. Wright TC, Stoler MH, Sharma A, Zhang G, Behrens C, Wright TL, ATHENA (Addressing THE Need for Advanced HPV Diagnostics) Study Group. 2011. Evaluation of HPV-16 and HPV-18 genotyping for the triage of women with high-risk HPV+ cytology-negative results. *Am J Clin Pathol* 136:578–586. <https://doi.org/10.1309/AJCPUS5EXAS6DKZ>.
25. Cicchetti DV, Feinstein AR. 1990. High agreement but low kappa: II. Resolving the paradoxes. *J Clin Epidemiol* 43:551–558. [https://doi.org/10.1016/0895-4356\(90\)90159-m](https://doi.org/10.1016/0895-4356(90)90159-m).
26. McHugh ML. 2012. Interrater reliability: the kappa statistic. *Biochem Med* 22:276–282. <https://doi.org/10.11613/BM.2012.031>.
27. Stanczuk G, Baxter G, Currie H, Lawrence J, Cuschieri K, Wilson A, Arbyn M. 2016. Clinical validation of hrHPV testing on vaginal and urine self-samples in primary cervical screening (cross-sectional results from the Papillomavirus Dumfries and Galloway–PaVDaG study). *BMJ Open* 6:e010660. <https://doi.org/10.1136/bmjopen-2015-010660>.
28. Sahasrabudhe VV, Gravitt PE, Dunn ST, Brown D, Allen RA, Eby YJ, Smith K, Zuna RE, Zhang RR, Gold MA, Schiffman M, Walker JL, Castle PE, Wentzensen N. 2014. Comparison of human papillomavirus detections in urine, vulvar, and cervical samples from women attending a colposcopy clinic. *J Clin Microbiol* 52:187–192. <https://doi.org/10.1128/JCM.01623-13>.
29. Stanczuk GA, Currie H, Baxter G, Foster A, Gibson L, Graham C, Cuschieri K. 2015. Cobas 4800 HPV detection in the cervical, vaginal and urine samples of women with high-grade CIN before and after treatment. *J Clin Pathol* 68:567–570. <https://doi.org/10.1136/clinpath-2014-202851>.
30. Pathak N, Dodds J, Zamora J, Khan K. 2014. Accuracy of urinary human papillomavirus testing for presence of cervical HPV: systematic review and meta-analysis. *BMJ* 349:g5264. <https://doi.org/10.1136/bmj.g5264>.
31. Giorgi-Rossi P, Franceschi S, Ronco G. 2012. HPV prevalence and accuracy of HPV testing to detect high-grade cervical intraepithelial neoplasia. *Int J Cancer* 130:1387–1394. <https://doi.org/10.1002/ijc.26147>.
32. Vorsters A, Van Damme P, Clifford G. 2014. Urine testing for HPV: rationale for using first void. *BMJ* 349:g6252. <https://doi.org/10.1136/bmj.g6252>.
33. Blake D, Crosbie E, Kitson S. 2017. Urinary HPV testing may offer hope for cervical screening non-attenders. *BJOG* 124:1364–1364. <https://doi.org/10.1111/1471-0528.14683>.