



Evaluation of the ResistancePlus MG FlexiBle Cartridge for Near-Point-of-Care Testing of *Mycoplasma genitalium* and Associated Macrolide Resistance Mutations

 E. L. Sweeney,^a L. P. Mhango,^a S. Ebeyan,^b L. Y. Tan,^b C. Bletchly,^c G. R. Nimmo,^{c,d} D. M. Whiley^{a,c}

^aUniversity of Queensland Centre for Clinical Research (UQ-CCR), Faculty of Medicine, The University of Queensland, Brisbane, Queensland, Australia

^bSpeedX Pty Ltd., Sydney, New South Wales, Australia

^cPathology Queensland Central Laboratory, Queensland, Australia

^dGriffith University, School of Medicine, Queensland, Australia

KEYWORDS *Mycoplasma genitalium*, SpeedX, near point-of-care, macrolide, azithromycin, antimicrobial resistance

Mycoplasma genitalium is an important sexually transmitted infection that can cause acute and/or chronic urethritis in males and females, and its prevalence is approximately 16% in females and 17% in males (1). Globally, macrolide resistance is estimated to exceed 50% in most urban centers (1–6), and in our local region of Queensland, Australia, it exceeds 60% (7). PCR is routinely used for diagnosis of *M. genitalium* and macrolide resistance mutations at positions 2058 and 2059 of the 23S rRNA gene; however, the turnaround times of laboratory-based methods often exceed 24 h and so may not be sufficiently timely to inform clinical management for symptomatic patients. Here, we evaluated the performance of the ResistancePlus MG FlexiBle cartridge test on the Cepheid GeneXpert (hereafter termed MG-Flex), which offers the potential for near-point-of-care testing.

A bank of 181 clinical samples (*M. genitalium* positive, $n = 63$, and negative, $n = 118$) from 145 males and 36 females was used in this study. Original patient samples were stored at 4°C for 4 weeks, then at –20°C for longer-term storage. Nucleic acid samples were stored at –20°C until required. Full details of the samples are provided in Table S1 in the supplemental material. The samples had all been submitted for routine *M. genitalium* testing to Pathology Queensland (Brisbane, Australia) where they were tested using an in-house PCR that detects the *MgPa* gene of *M. genitalium* (hereafter termed in-house *MgPa*-PCR) (8). For the purposes of this evaluation, all samples were tested using the MG-Flex near-point-of-care assay (which detects the macrolide resistance mutations A2058T, A2058C, A2058G, and A2059G) as well as the SpeedX ResistancePlus MG test (hereafter termed RPMG, which detects the macrolide resistance mutations A2058T, A2058C, A2058G, A2059C, and A2059G). The RPMG test is a Therapeutic Goods Administration (TGA, Australia) cleared and CE-IVD-marked (i.e., Conformité Européenne-marked for in vitro diagnostic) laboratory-based test.

The MG-Flex assay was prepared per the kit instructions. Briefly, 44 μ l of MG-Flex mastermix, 1 ml of neat (unextracted) clinical sample, and 10 μ l of internal control were added to the FlexiBle cartridge (Cepheid). In some instances, swab specimens ($n = 17$; Table S1) had less than 1 ml of sample, so the volume was made up to 1 ml with sterile molecular-grade water. The RPMG assay was performed per the manufacturer's instructions using stored DNA extracts from the routine testing in the in-house *MgPa*-PCR.

This study was approved by the Children's Health Queensland human research ethics committee (HREC/12/QRCH/139).

The results are summarized in Table 1 and further detailed in Table S1. The MG-Flex assay detected *M. genitalium* in 61 of 63 known positive samples, providing a sensitivity

Citation Sweeney EL, Mhango LP, Ebeyan S, Tan LY, Bletchly C, Nimmo GR, Whiley DM. 2020. Evaluation of the ResistancePlus MG FlexiBle cartridge for near-point-of-care testing of *Mycoplasma genitalium* and associated macrolide resistance mutations. J Clin Microbiol 58:e01897-19. <https://doi.org/10.1128/JCM.01897-19>.

Editor Erik Munson, Marquette University

Copyright © 2020 American Society for Microbiology. All Rights Reserved.

Address correspondence to E. L. Sweeney, e.l.sweeney@uq.edu.au, or D. M. Whiley, d.whiley@uq.edu.au.

Accepted manuscript posted online 2 January 2020

Published 24 February 2020

TABLE 1 Summary of results of the RPMG and MG-Flex assays^a

No. of samples	Gender	Sample type (no.)	In-house MgPa-PCR result	RPMG assay result	MG-Flex assay result
7	F	Cervical (3), vaginal (1), urine (3)	Detected	MG detected. 23S rRNA mutations detected.	MG detected. 23S rRNA mutations detected.
36	M	Urine (31), rectal (3), urethral (2)	Detected	MG detected. 23S rRNA mutations detected.	MG detected. 23S rRNA mutations detected.
5	F	Cervical (2), vaginal (2), urine (1)	Detected	MG detected. 23S rRNA mutations not detected.	MG detected. 23S rRNA mutations not detected.
13	M	Urine (11), rectal swab (2)	Detected	MG detected. 23S rRNA mutations not detected.	MG detected. 23S rRNA mutations not detected.
2	M	Urine (1), rectal swab (1)	Detected	MG detected. 23S rRNA mutations not detected.	MG not detected. 23S rRNA mutations not detected.
92	M	Urine (85), rectal swab (6), genital swab (1)	ND	MG not detected. 23S rRNA mutations not detected.	MG not detected. 23S rRNA mutations not detected.
24	F	Cervical (4), vaginal (2), urine (16), rectal (2)	ND	MG not detected. 23S rRNA mutations not detected.	MG not detected. 23S rRNA mutations not detected.
2	M	Rectal swab (1), urine (1)	ND	MG not detected. 23S rRNA mutations not detected.	Invalid

^aBold text represents a significant result (i.e., *M. genitalium* was detected or a result was invalid). F, female; M, male; ND, not detected.

of 96.8%. The two samples providing negative results by the MG-Flex assay (samples 62 and 63; Table S1) were both from males. For sample 62, a reduced starting volume of sample likely contributed to discordance, and in both samples, low *M. genitalium* load was observed. One hundred sixteen *M. genitalium*-negative samples were negative in the MG-Flex assay, while two samples provided invalid results (samples 180 and 181; Table S1), and there was insufficient sample to repeat testing. MG-Flex assay specificity for evaluable samples was therefore 100%. The detection of macrolide resistance mutations by the MG-Flex assay correlated 100% with that of the RPMG assay, with the exception of two samples providing negative results for *M. genitalium* in the MG-Flex assay (samples 62 and 63; Table S1). The overall agreement between the RPMG and MG-Flex assays was 98.9%, with a kappa value of 0.98. The detection limit was also assessed by testing 10-fold dilutions of an *M. genitalium*-positive sample, and the MG-Flex test was able to reliably detect an additional dilution over the RPMG assay (Table S1).

In summary, we found that the MG-Flex cartridge test was highly sensitive and specific for the detection of *M. genitalium* and 23S rRNA mutations.

SUPPLEMENTAL MATERIAL

Supplemental material is available online only.

SUPPLEMENTAL FILE 1, PDF file, 0.5 MB.

ACKNOWLEDGMENTS

We thank Pathology Queensland staff for their assistance with this study.

D.M.W. reports research funding from SpeeDx Pty Ltd. The RPMG and MG-Flex assays were provided by SpeeDx Pty Ltd. Co-authors from SpeeDx Pty Ltd. (S.E. and L.Y.T.) participated in the study conception and design, data analysis, and manuscript writing.

This work was supported by The University of Queensland strategic funding.

REFERENCES

1. Getman D, Jiang A, O'Donnell M, Cohen S. 2016. Mycoplasma genitalium prevalence, coinfection, and macrolide antibiotic resistance frequency in a multicenter clinical study cohort in the United States. *J Clin Microbiol* 54:2278–2283. <https://doi.org/10.1128/JCM.01053-16>.
2. Dumke R, Thurmer A, Jacobs E. 2016. Emergence of Mycoplasma genitalium strains showing mutations associated with macrolide and fluoroquinolone resistance in the region Dresden, Germany. *Diagn Microbiol Infect Dis* 86: 221–223. <https://doi.org/10.1016/j.diagmicrobio.2016.07.005>.
3. Gesink D, Racey CS, Seah C, Zittermann S, Mitterni L, Juzkiw J, Jamieson H, Greer J, Singh S, Jensen JS, Allen V. 2016. Mycoplasma genitalium in Toronto, Ont: estimates of prevalence and macrolide resistance. *Can Fam Physician* 62:e96–e101.
4. Jensen JS, Bradshaw CS, Tabrizi SN, Fairley CK, Hamasuna R. 2008. Azithromycin treatment failure in Mycoplasma genitalium-positive patients with nongonococcal urethritis is associated with induced macrolide resistance. *Clin Infect Dis* 47:1546–1553. <https://doi.org/10.1086/593188>.
5. Gratrix J, Plitt S, Turnbull L, Smyczek P, Brandley J, Scarrott R, Naidu P, Parker P, Blore B, Bull A, Shokoples S, Bertholet L, Martin I, Chernesky M, Read R, Singh A. 2017. Prevalence and antibiotic resistance of Mycoplasma genitalium among STI clinic attendees in Western Canada: a cross-sectional analysis. *BMJ Open* 7:e016300. <https://doi.org/10.1136/bmjopen-2017-016300>.
6. Deguchi T, Yasuda M, Horie K, Seike K, Kikuchi M, Mizutani K, Tsuchiya T, Yokoi S, Nakano M, Hoshina S. 2015. Drug resistance-associated mutations in Mycoplasma genitalium in female sex workers, Japan. *Emerg Infect Dis* 21:1062–1064. <https://doi.org/10.3201/eid2106.142013>.
7. Sweeney EL, Trembizki E, Bletchly C, Bradshaw CS, Menon A, Francis F, Langton-Lockton J, Nimmo GR, Whiley DM, Sweeney EL, Trembizki E, Bletchly C, Bradshaw CS, Menon A, Francis F, Langton-Lockton J, Nimmo GR, Whiley DM. 2019. Levels of Mycoplasma genitalium antimicrobial resistance differ by both region and gender in the state of Queensland, Australia: implications for treatment guidelines. *J Clin Microbiol* 57. <https://doi.org/10.1128/JCM.01555-18>.
8. Jensen JS, Bjornelius E, Dohn B, Lidbrink P. 2004. Use of TaqMan 5' nuclease real-time PCR for quantitative detection of Mycoplasma genitalium DNA in males with and without urethritis who were attendees at a sexually transmitted disease clinic. *J Clin Microbiol* 42:683–692. <https://doi.org/10.1128/jcm.42.2.683-692.2004>.